

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: Thermo

Model: TSX0086D

Serial No: 1124092101210920

Department/Location: Delp G040

Set Point (°C): -80°C

Certified Temperature (°C): -79.4°C

Observed Temperature (°C): -79°C

Observed Accuracy (°C): 0.4°C

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II

Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm

Date: 1/12/23

(Print)

Service Technician: 

(Sign)

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: Fisher

Model: ISOTEMP

Serial No: A434F1FF2A93

Department/Location: Sudler 2012E

Asset#: S-4

Set Point (°C): N/A

Certified Temperature (°C): 2.6°c

Observed Temperature (°C): N/A

Observed Accuracy (°C): 0°c

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II


Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: AEGIS

Model: 2-F-4

Serial No: 0820872-38728-Y

Department/Location: Sudler 2012E

Asset#: S-3

Set Point (°C): -20°c

Certified Temperature (°C): -19.6°c

Observed Temperature (°C): -19.6°c

Observed Accuracy (°C): 0°c

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II


Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: American Biotech Supply (ABS)

Model: ABT-HC-19-TS

Serial No: ABS-9576207-1907

Department/Location: Sudler 2012E

Asset#: S-2

Set Point (°C): 5 °c

Certified Temperature (°C): 4.7°c

Observed Temperature (°C): 4.5°c

Observed Accuracy (°C): 0.2°c

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II

Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: Thermo

Model: TSX60086A

Serial No: 1117686301191108

Department/Location: Sudler 2012E

Asset#: S-1

Set Point (°C): -80°C

Certified Temperature (°C): -78°C

Observed Temperature (°C): -78.1°C

Observed Accuracy (°C): 0.1°C

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II


Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: Accucold

Model: VT65ML

Serial No: 2021 12 000719

Department/Location: Delp G040

Set Point (°C): -25°c

Certified Temperature (°C): -24.5°c

Observed Temperature (°C): -24.4°c

Observed Accuracy (°C): 0.1°c

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II

Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: Thermo

Model: 3751FS

Serial No: 0146956301121018

Department/Location: Delp G040

Set Point (°C): N/A

Certified Temperature (°C): N/A

Observed Temperature (°C): -15°c

Observed Accuracy (°C): 0°c

Passed/Failed: Pass

Certification Date: 1/12/23

Certification Due: 1/2024

Certifying Equipment

Manufacturer: Fluke

Model: 54II

Serial No: 80670027

Due Date: 07/8/2023

Service Technician: Scott Hamm
(Print)

Date: 1/12/23

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Preventive Maintenance

Customer Information

Company Name: University of KS Health Syst-CTSU Rainbow

Company Address: 3901 Rainbow Blvd. Kansas City, MO 66160

Contact Person: Shelby Williams 913-574-3018

Thermometer Information: Built-in

Manufacturer: ICE- O - MATIC

Model: GEMD270A

Serial No: Y9 171480 18P

Department/Location: Sudler 2012E

Asset# S-5

CLEANED CONDENSER

CHECKED OPERATION

Service Technician: Scott Hamm

Date: 1/12/23

(Print)

Service Technician: 

(Sign)