

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Scale Performance Report

Customer Information

Company Name: KU Med Center

Company Address: 3901 Rainbow, Kansas City MO

Contact Person/ Phone Number: Shelby Williams, 913-574-3018,
swilliams46@kumc.edu

Manufacturer: Health-o-Meter

Model: 500KL

Serial No: 500 0163137

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Operation:	Pass	X	Fail

Operation:

Measured Values	Set:	50lbs	Measured:	49.6lbs
	Set:	250lbs	Measured:	251.0lbs
	Set:	500lbs	Measured:	501.8lbs

Certification Date: 1/11/2024

Certification Due: 1/2025

Certifying Equipment

Kansas Metrology Laboratory Weight Set Certificate Number: K16924 Date Due: 4/3/24

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Nick Brown

Date: 1/11/24

Service Technician: _____
(Sign)