

**COBB**  
REFRIGERATION  
A TEMP-CON COMPANY

# CERTIFICATION OF ACCURACY

## Customer Information

**Company Name:** KUMC Swope CTSU

**Company Address:** 4001 Dr. Martin Luther King Jr. Blvd, Kansas City, MO 64130

**Contact Person/ Phone Number:** Autumn Beauchamp [abeauchamp@kumc.edu](mailto:abeauchamp@kumc.edu)

---

**Manufacturer:** Burdick

**Model:** ELI 280

**Serial No:** 122380001963

**ECG Setting:** 60 bpm

**ECG Reading:** 60bpm

**Passed/Failed:** Pass

**Certification Date:** 1/12/2026

**Certification Due:** 1/2027

---

## Certifying Equipment

**Manufacturer:** Fluke

**Model:** Prosim8

**Serial No:** 3990033

---

**Service Technician:** Derek Hulett  
(Print)

**Date:** 1/12/2026

**Service Technician:** Derek Hulett  
(Sign)