

**COBB Refrigeration & Laboratory Services LLC**

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

**CERTIFICATION OF ACCURACY**

**Customer Information**

**Company Name: University of KS Medical Center-CTSU**

**Company Address: 3901 Rainbow Blvd Kansas City, KS 66103**

**Contact Person: Shelby Williams**

**Phone Number: 913-574-3018**

**Thermometer Information: Built-In**

**Manufacturer: Fisher Scientific**

**Model: 3751FS**

**Serial No: 0146956301121018**

**Department/Location: CTSU Rainbow**

**Asset # N/A**

**Set Point: N/A**

**Certified Temperature: -2.2°C**

**Observed Temperature: N/A**

**Elemental Monitor: -3.07°C**

**Observed Accuracy: 0.87°C**

**Passed/Failed: Pass**

**Certification Date: 1/10/2024**

**Certification Due: 1/10/2025**

**Certifying Equipment**

**Manufacturer: Fluke**

**Model: 54iiB**

**Serial No: 86470112**

**Due Date: 04/18/2024**

**Service Technician: Nicholas Brown  
(Print)**

**Date: 1/10/2024**

**Service Technician:  
(Sign)**

