

**Thermometer Performance Report(ORAL)**

Customer:		University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)					Contact:		Nelda Mallett						
Street:		3901 Rainbow					Phone #		913-574-3007						
City/ST/Zip:		Kansas City, KS 66160					+/- 2 (C)=Pass		P= Pass F=Fail						
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Performed	Date Due	
											Unit Temp	Certified Temp			
NA	901053	WelchAllyn	(21)19147463	P	P	P	P	P	P	P	37.0	37.1	1/25/2023	1/2024	
NA	690	WelchAllyn	04293852	P	P	P	P	P	P	P	37.0	37.0	1/25/2023	1/2024	
NA	901053	WelchAllyn	(21)19147495	P	P	P	P	P	P	P	37.0	37.0	1/13/2023	1/2024	
NA	901053	WelchAllyn	(21)19147359	P	P	P	P	P	P	P	37.1	37.0	1/13/2023	1/2024	
NA	0123	Covidien	A1952323x	P	P	P	P	P	P	P	37.0	37.0	1/25/2023	1/2024	
NA	0123	Covidien	A1952353X	P	P	P	P	P	P	P	37.0	37.0	1/13/2023	1/2024	
NA	0123	Covidien	SMF0812-001	P	P	P	P	P	P	P	37.0	37.0	1/25/2023	1/2024	

Technician:

Test equipment used  
 FLUKE- 52II Thermometer SN: 45470242WS EX: 07/12/2023,  
 Certificate# 00074293

Notes:

**Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088**

Thermometer Performance Report(ORAL)														
Customer:		University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)				Contact:		Nelda Mallett						
Street:		3901 Rainbow				Phone #		913-574-3007						
City/ST/Zip:		Kansas City, KS 66160				+/- 2 (C)=Pass		P= Pass		F=Fail				
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Peformed	Date Due
											Unit Temp	Certified Temp		

Thermometer Performance Report(ORAL)														
Customer:		University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)				Contact:		Nelda Mallett						
Street:		3901 Rainbow				Phone #		913-574-3007						
City/ST/Zip:		Kansas City, KS 66160				+/- 2 (C)=Pass		P= Pass		F=Fail				
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Peformed	Date Due
											Unit Temp	Certified Temp		