

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center-CTSU

Company Address: 3901 Rainbow Blvd Kansas City, KS 66103

Contact Person: Shelby Williams

Phone Number: 913-574-3018

Thermometer Information: Built-In

Manufacturer: ABS American Biotech Supply

Model: ABT-HC-19-TS

Serial No: ABS-9576207-1907

Department/Location: CTSU Rainbow

Asset # N/A

Set Point: 4°C

Certified Temperature: 4°C

Observed Temperature: 4.1°C

Elemental Monitor: 3.87°C

Observed Accuracy: .13°C

Passed/Failed: Pass

Certification Date: 1/10/2024

Certification Due: 1/10/2025

Certifying Equipment

Manufacturer: Fluke

Model: 54iiB

Serial No: 86470112

Due Date: 04/18/2024

Service Technician: Nicholas Brown
(Print)

Date: 1/10/2024

Service Technician: 
(Sign)