COBB Refrigeration LLC

913-403-0088 • www.cobbrls.com

Refrigeration, Freezer Performance Report

Customer:	UKMC-Swope CTSU	Contact:	Tricia Snow
Street:	4000 Dr. Martin Luther King Jr. Blvd	Phone #	816-599-5310
City/ST/Zip:	Kansas City, MO 64130	+/- 2 C=Pass	P= Pass F=Fail

Unit	Model#	MFG	S/N	Bat#/Date Rep	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Calibration Pass/Fail +/- 2C	Temperature (Celsius) Certified Temp	Temperature (Celsius) Unit Temp	Date Performed
REF	GTFBV05CPSA	FISHER	300422763	NA	P	P	P	P	2.2	2.6	11/14/2022
FRZ	GTFBV05CPSA	FISHER	300422763	NA	P	P	P	P	-10.3	-9.5	11/14/2022
Ultra-Low	RDE40086FA	THERMO	1127312701220719	NO	P	P	P	P	-80.2	-80	11/14/2022
FRZ	253.27002410	KENMORE	WB52350497	NA	P	P	P	P	-28.4	-27.7	11/14/2022

Technicians Initials: JC

Technician: Jim Cowart

Date: 11/14/2022

Test equipment used: FLUKE- 52II Thermometer SN: 45470242WS EX: 07/12/2023

Notes:

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St.. Lenexa KS, 66212 913-403-0088