

# Service Report



## WORK ORDER ID: CWKD4082828

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy. Ste. #1360 Fairway, KS 66205

AGILITI ASSET ID : 1887916 | SERIAL #: N/A | VENDOR SITE ID: | CUSTOMER ASSET ID: 412269-86

ASSET TYPE: FREEZERS, LABORATORY | MANUFACTURER: THERMO SCIENTIFIC | MODEL: THERMO SCIENTIFIC LAB FREEZER

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887916

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

### LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Kirstin Williams	PM Not Covered	0.5	M-F 8-5	01/23/24

### SERVICE PROVIDED DETAILS

Performed SM per OEM specifications. All test passed.

### TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
DM1	459	DIGITAL MULTIMETER	04/17/23
QC8	139	SAFETY ANALYZER	04/14/23

**CHECKLIST**

**PROCEDURE: L1100 L1**

<b>QUESTION</b>	<b>ANSWER</b>
CES THERM. - FREEZER (if applicable)	-73
CES THERM. - REFRIGERATOR	N/A
CUST. THERM. - FREEZER (if applicable)	-73
CUST. THERM. - REFRIGERATOR	N/A
DESIRED TEMP. - FREEZER (if applicable)	-73
DESIRED TEMP. - REFRIGERATOR	N/A
Electrical Safety Inspection	Pass
GASKET SEAL	Pass
High Temperature Alarm	Yes
Low Temperature Alarm	Yes
Mechanical Safety Inspection	Pass
Operational Check	Pass
STANDS LEVEL	Pass
STANDS SOLIDLY	Pass

---

Customer Signature:

Date:

---

Technician Signature:

Date: