

AED/Power Vac Performance Report														
Customer: University of KS Health Systems CTSU				Contact: Shelby Williams										
Street: 4350 Shawnee Mission Pkwy				Phone #: 913-574-3018										
City/ST/Zip: Fairway, KS 66205				P= Pass F=Fail										
Asset#	Model#	MFG	S/N	Pad Ex. Date	CPR-D Padz/EX-Date	Condition Controls/Switches	Manual Self Test	Hardware Integrity	Indicators & Display Operation	MFG. Labeling Intact	Batteries Replaced Y/N	Date Due	Calibration Performed	Calibration Due
034856	197	Zoll	AA128023827	2026-16-04	P	P	P	P	P	P	N	EX-03/2025	1/31/2022	01/2023
034857	197	Zoll	AA128023825	2026-16-04	P	P	P	P	P	P	N	EX-03/2025	1/31/2022	01/2023
<b>Power Vacume</b>				<25 inHgVac										
033860	PM61	Power vac	002535	Yes		P	P	P	P	P	NA	NA	1/31/2022	01/2023
033844	PM61	Power vac	002538	Yes		P	P	P	P	P	NA	NA	1/31/2022	01/2023

Technicians	Jim Cowart JC	Technician:	Jim Cowart	Test equipment used	
Initials					
Notes:					