

# Service Report



## WORK ORDER ID: CWKD4082865

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy, Ste. #1360 Fairway, KS 66205

AGILITI ASSET ID : 1887912 | SERIAL #: 7075159 | VENDOR SITE ID: | CUSTOMER ASSET ID:

ASSET TYPE: FREEZERS, LABORATORY | MANUFACTURER: AEGIS SCIENTIFIC | MODEL: SERIES II FREEZER

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887912

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

### LABOR

| SERVICE PROVIDER | CATEGORY       | HOURS | PERIOD DEFINITION | SERVICE DATE |
|------------------|----------------|-------|-------------------|--------------|
| Kirstin Williams | PM Not Covered | 0.5   | M-F 8-5           | 01/23/24     |

### SERVICE PROVIDED DETAILS

Performed SM per OEM specifications. All test passed.

### TEST EQUIPMENT

| PREFIX | UNIT | DESCRIPTION        | CALIBRATION DATE |
|--------|------|--------------------|------------------|
| QC8    | 139  | SAFETY ANALYZER    | 04/14/23         |
| DM1    | 459  | DIGITAL MULTIMETER | 04/17/23         |

**CHECKLIST**

**PROCEDURE: L1100 L1**

**QUESTION**

**ANSWER**

CES THERM. - FREEZER (if applicable)

-10

CES THERM. - REFRIGERATOR

N/A

CUST. THERM. - FREEZER (if applicable)

-9

CUST. THERM. - REFRIGERATOR

N/A

DESIRED TEMP. - FREEZER (if applicable)

-15.8

DESIRED TEMP. - REFRIGERATOR

N/A

Electrical Safety Inspection

Pass

GASKET SEAL

Pass

High Temperature Alarm

Yes

Low Temperature Alarm

Yes

Mechanical Safety Inspection

Pass

Operational Check

Pass

STANDS LEVEL

Pass

STANDS SOLIDLY

Pass

---

Customer Signature:

Date:

---

Technician Signature:

Date: