

Service Report



WORK ORDER ID: CWKD4082978

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy. Ste. #1360 Fairway, KS 66205

AGILITY ASSET ID : 1887913 | SERIAL #: 192423688 | VENDOR SITE ID: | CUSTOMER ASSET ID:

ASSET TYPE: INCUBATORS, LABORATORY, AEROBIC | MANUFACTURER: BOEKEL SCIENTIFIC INC | MODEL: DIGITAL INCUBATOR

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887913

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Kirstin Williams	PM Not Covered	0.5	M-F 8-5	01/23/24

SERVICE PROVIDED DETAILS

Performed SM per OEM specifications. All test passed.

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
DM1	459	DIGITAL MULTIMETER	04/17/23

CHECKLIST

PROCEDURE: L1200 L1

QUESTION

- Alarm - High Temperature
- Alarm - Low Temperature
- CES THERM.
- Clinical- Audio
- Clinical- Visual
- CUST. THERM.
- DESIRED TEMP
- Electrical Safety Inspection
- Mechanical Safety Inspection
- Operational Check
- Patient- Audio
- Patient- Visual
- Remote
- Safety- Audio
- Safety- Visual

ANSWER

- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass
- Pass

Customer Signature:

Date:

Technician Signature:

Date: