

Sudler 2012 CTSU Lab Fridge
for
Sudler 2012 CTSU Lab Fridge

Jan 01, 2024 - Feb 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Lab Fridge" for Sudler 2012 CTSU Lab Fridge

Jan 01, 2024 - Feb 01, 2024

MACHINE DETAILS


Serial Number: **0081F90EC68E**
Location: **CTSU Rainbow Lab**
Description: **0081F90EC68E**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **4.23 °C** | MIN: **-0.07 °C** | MAX: **11.77 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2023-12-31 6:00 AM	2.14°C
2024-01-01 6:00 AM	5.72°C
2024-01-02 6:00 AM	5.16°C
2024-01-03 6:00 AM	5.74°C
2024-01-04 6:00 AM	4.82°C
2024-01-05 6:00 AM	3.8°C
2024-01-06 6:00 AM	5.1°C
2024-01-07 6:00 AM	5.24°C
2024-01-08 6:00 AM	5.08°C
2024-01-09 6:00 AM	3.22°C
2024-01-10 6:00 AM	2.67°C
2024-01-11 6:00 AM	1.2°C
2024-01-12 6:00 AM	2.34°C
2024-01-13 6:00 AM	5.83°C
2024-01-14 6:00 AM	1.47°C
2024-01-15 6:00 AM	5.54°C
2024-01-16 6:00 AM	3.94°C
2024-01-17 6:00 AM	3.86°C
2024-01-18 6:00 AM	2.38°C
2024-01-19 6:00 AM	3.23°C
2024-01-20 6:00 AM	1.95°C
2024-01-21 6:00 AM	4.87°C
2024-01-22 6:00 AM	0.47°C
2024-01-23 6:00 AM	5.35°C
2024-01-24 6:00 AM	3.81°C
2024-01-25 6:00 AM	2.6°C
2024-01-26 6:00 AM	4.52°C
2024-01-27 6:00 AM	3.86°C
2024-01-28 6:00 AM	3.06°C
2024-01-29 6:00 AM	5.55°C
2024-01-30 6:00 AM	2.16°C

CENTRAL TIME (US & CANADA)

TEMPERATURE (°C)

2024-01-31 6:00 AM

5.41°C

ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Lab Fridge
for
Sudler 2012 CTSU Lab Fridge

Feb 01, 2024 - Mar 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Lab Fridge" for Sudler 2012 CTSU Lab Fridge

Feb 01, 2024 - Mar 01, 2024

MACHINE DETAILS


Serial Number: 0081F90EC68E
Location: CTSU Rainbow Lab
Description: 0081F90EC68E
Installation date:

EQUIPMENT DETAILS

Category: +4 Refrigerator
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: 4.34 °C | MIN: -0.84 °C | MAX: 13.12 °C



Timezone: Central Time (US & Canada) (CST)

● Temperature (°C)

DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-01-31 6:00 AM	5.87 °C
2024-02-01 6:00 AM	4.35 °C
2024-02-02 6:00 AM	2.47 °C
2024-02-03 6:00 AM	2.29 °C
2024-02-04 6:00 AM	5.59 °C
2024-02-05 6:00 AM	3.69 °C
2024-02-06 6:00 AM	2.8 °C
2024-02-07 6:00 AM	4.77 °C
2024-02-08 6:00 AM	4.66 °C
2024-02-09 6:00 AM	5.48 °C
2024-02-10 6:00 AM	3.6 °C
2024-02-11 6:00 AM	5.63 °C
2024-02-12 6:00 AM	2.34 °C
2024-02-13 6:00 AM	2.06 °C
2024-02-14 6:00 AM	5.98 °C
2024-02-15 6:00 AM	11.4 °C
2024-02-16 6:00 AM	3.65 °C
2024-02-17 6:00 AM	3.27 °C
2024-02-18 6:00 AM	5.6 °C
2024-02-19 6:00 AM	3.6 °C
2024-02-20 6:00 AM	5.32 °C
2024-02-21 6:00 AM	1.85 °C
2024-02-22 6:00 AM	5.77 °C
2024-02-23 6:00 AM	5.74 °C
2024-02-24 6:00 AM	5.49 °C
2024-02-25 6:00 AM	3.3 °C
2024-02-26 6:00 AM	3.48 °C
2024-02-27 6:00 AM	1.69 °C
2024-02-28 6:00 AM	5.89 °C
2024-02-29 6:00 AM	2.43 °C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-29 4:01 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.57 °C	8.0/2.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-29 4:01 AM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.57 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-02-29 4:01 AM Message:Temperature Clear: 2.57 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-29 3:49 AM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.95 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-02-29 3:49 AM Message:Temperature Alert: 0.95 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-02-28 11:10 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.45 °C	8.0/2.0 °C	Alert Cleared
						<p>● 2024-02-28 11:10 PM Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.45 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-02-28 11:10 PM Message:Temperature Clear: 2.45 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-28 10:53 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.23 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-02-28 10:53 PM Message:Temperature Alert: 1.23 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-28 10:29 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.14 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-02-28 10:29 PM Message:Temperature Alert: 1.14 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-02-28 5:43 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	4.02 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-28 3:35 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.62 °C	8.0/2.0 °C	Alert Cleared
2024-02-28 5:04 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.08 °C	8.0/2.0 °C	Alert Cleared




CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-27 10:05 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.12 °C	8.0/2.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.12 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Clear: 2.12 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.80 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Alert: 1.80 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.24 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Alert: 1.24 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.45 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Alert: 1.45 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-02-27 6:49 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.48 °C	8.0/2.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.48 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Clear: 2.48 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.70 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Message:Temperature Alert: 1.70 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-02-27 5:23 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.15 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 5:23 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.15 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-27 5:23 AM					Message:Temperature Clear: 2.15 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 4:52 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.12 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-27 4:52 AM					Message:Temperature Alert: 1.12 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 4:28 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.97 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-27 4:28 AM					Message:Temperature Alert: 0.97 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 4:04 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.37 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-27 4:04 AM					Message:Temperature Alert: 1.37 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
2024-02-27 4:34 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.71 °C	8.0/1.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 4:34 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.71 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
●	2024-02-27 4:34 AM					Message:Temperature Clear: 1.71 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
●	2024-02-27 4:28 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
●	2024-02-27 4:28 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-02-27 2:58 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.45 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.45 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 2:58 AM					
						Message:Temperature Clear: 2.45 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 2:58 AM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.02 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 2:34 AM					
						Message:Temperature Alert: 1.02 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 2:34 AM					
2024-02-27 1:06 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	5.45 °C	8.0/2.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 5.45 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 1:06 AM					
						Message:Temperature Clear: 5.45 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 1:06 AM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.67 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:52 AM					
						Message:Temperature Alert: 1.67 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:52 AM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.16 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:27 AM					
						Message:Temperature Alert: 1.16 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:27 AM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.59 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:03 AM					
						Message:Temperature Alert: 0.59 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-27 12:03 AM					
2024-02-26 9:58 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.91 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 9:58 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.91 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:Delp G040 +4</p> <p>● 2024-02-26 9:58 PM Message:Temperature Clear: 3.91 °C, within the limit of 1.0 °C to 8.0 °C</p> <p>Alert Rule:Delp G040 +4</p>						
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 9:38 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.67 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:Delp G040 +4</p> <p>● 2024-02-26 9:38 PM Message:Temperature Alert: 0.67 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p>						
2024-02-26 5:59 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.53 °C	8.0/2.0 °C	Alert Cleared
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 5:59 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.53 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>● 2024-02-26 5:59 PM Message:Temperature Clear: 2.53 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>						
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 5:47 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.77 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>● 2024-02-26 5:47 PM Message:Temperature Alert: 0.77 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>						
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 5:23 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.95 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>● 2024-02-26 5:23 PM Message:Temperature Alert: 1.95 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>						
<p>● Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-26 4:58 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.26 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>● 2024-02-26 4:58 PM Message:Temperature Alert: 1.26 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>						
2024-02-26 5:53 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.16 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 5:53 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.16 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.16 °C, within the limit of 1.0 °C to 8.0 °C
	2024-02-26 5:53 PM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 5:47 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.77 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.77 °C, below the limit of 1.0 °C
	2024-02-26 5:47 PM					Alert Rule:Delp G040 +4
2024-02-26 4:46 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.41 °C	8.0/1.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 4:46 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.41 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 2.41 °C, within the limit of 1.0 °C to 8.0 °C
	2024-02-26 4:46 PM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 4:39 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.94 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.94 °C, below the limit of 1.0 °C
	2024-02-26 4:39 PM					Alert Rule:Delp G040 +4
2024-02-26 11:43 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	4.7 °C	8.0/2.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 11:43 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.70 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 4.70 °C, within the limit of 2.0 °C to 8.0 °C
	2024-02-26 11:43 AM					Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-26 11:36 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.86 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.86 °C, below the limit of 2.0 °C
	2024-02-26 11:36 AM					Alert Rule:SUDLER 2012 CTSU lab Fridge

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-26 9:24 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.52 °C	8.0/2.0 °C	Alert Cleared
	2024-02-26 9:24 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.52 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 9:24 AM	Message:Temperature Clear: 2.52 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 9:00 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.69 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 9:00 AM	Message:Temperature Alert: 0.69 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
2024-02-26 6:12 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	3.28 °C	8.0/2.0 °C	Alert Cleared
	2024-02-26 6:12 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.28 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 6:12 AM	Message:Temperature Clear: 3.28 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 6:05 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.73 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 6:05 AM	Message:Temperature Alert: 1.73 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 5:40 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.04 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-02-26 5:40 AM	Message:Temperature Alert: 1.04 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
2024-02-25 9:43 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.53 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-25 2:30 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.41 °C	8.0/2.0 °C	Alert Cleared
2024-02-25 12:05 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.33 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
●	2024-02-25 12:05 PM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.33 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 12:05 PM					Message:Temperature Clear: 2.33 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 11:45 AM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.28 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 11:45 AM					Message:Temperature Alert: 1.28 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 11:21 AM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.41 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 11:21 AM					Message:Temperature Alert: 1.41 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-02-25 10:05 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.41 °C	8.0/2.0 °C	Alert Cleared
---------------------	-----------------------------	-------------	------------------------------------	---------	------------	---------------

●	2024-02-25 10:05 AM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.41 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 10:05 AM					Message:Temperature Clear: 2.41 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

●	2024-02-25 9:46 AM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.63 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 9:46 AM					Message:Temperature Alert: 1.63 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

●	2024-02-25 9:20 AM					Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.50 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-02-25 9:20 AM					Message:Temperature Alert: 0.50 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-02-25 8:49 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.37 °C	8.0/1.0 °C	Alert Cleared
--------------------	-----------------------------	-------------	---------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.37 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.37 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.67 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.67 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-02-25 2:52 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.87 °C 8.0/2.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.87 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 2.87 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.00 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.00 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.84 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.84 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.84 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 0.84 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-02-24 9:02 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 3.27 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-24 7:30 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.49 °C	8.0/2.0 °C	Alert Cleared
	2024-02-24 7:30 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.49 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 7:30 AM		Message:Temperature Clear: 2.49 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:52 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.46 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:52 AM		Message:Temperature Alert: 0.46 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:28 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.47 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:28 AM		Message:Temperature Alert: 1.47 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:02 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.47 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 6:02 AM		Message:Temperature Alert: 1.47 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-24 4:56 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	3.91 °C	8.0/2.0 °C	Alert Cleared
	2024-02-24 4:56 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.91 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 4:56 AM		Message:Temperature Clear: 3.91 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 4:49 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.89 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-24 4:49 AM		Message:Temperature Alert: 1.89 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-24 3:29 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.41 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-24 12:38 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.06 °C	8.0/2.0 °C	Alert Cleared
2024-02-24 12:14 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.41 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-24 12:14 AM		●	Notification:Email, SMS, and App sent to Miranda Wright.			
		●	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.41 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
2024-02-24 12:14 AM		●	Message:Temperature Clear: 2.41 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			
		●	Notification:Email, SMS, and App sent to Miranda Wright.			
2024-02-24 12:07 AM		●	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
		●	Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			
2024-02-23 10:13 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.45 °C	8.0/1.0 °C	Alert Cleared
2024-02-23 10:13 PM		●	Notification:Email, SMS, and App sent to Miranda Wright.			
		●	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.45 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
2024-02-23 10:13 PM		●	Message:Temperature Clear: 1.45 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			
		●	Notification:Email, SMS, and App sent to Miranda Wright.			
2024-02-23 10:07 PM		●	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.56 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
		●	Message:Temperature Alert: 0.56 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			
2024-02-23 6:22 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.34 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.34 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 6:22 PM					
						Message:Temperature Clear: 2.34 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 6:22 PM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.73 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:57 PM					
						Message:Temperature Alert: 1.73 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:57 PM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.76 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:34 PM					
						Message:Temperature Alert: 1.76 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:34 PM					
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.94 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:09 PM					
						Message:Temperature Alert: 1.94 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-23 5:09 PM					
2024-02-23 3:59 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.45 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-23 3:59 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:47 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:47 PM	Sudler 2012 CTSU Lab Fridge	Message	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:23 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:23 PM	Sudler 2012 CTSU Lab Fridge	Message	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:00 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:00 PM	Sudler 2012 CTSU Lab Fridge	Message	Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-02-23 3:36 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.56 °C	8.0/1.0 °C	Alert Cleared
2024-02-23 3:36 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:Delp G040 +4			
2024-02-23 3:29 PM	Sudler 2012 CTSU Lab Fridge	Notification	Alert Rule:Delp G040 +4			
2024-02-23 3:29 PM	Sudler 2012 CTSU Lab Fridge	Message	Alert Rule:Delp G040 +4			
2024-02-23 3:08 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	4.73 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.73 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 4.73 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.61 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.61 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-02-23 1:45 PM Sudler 2012 CTSU Lab Fridge Temperature **Delp G040 +4** 1.52 °C 8.0/1.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.52 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.52 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.71 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.71 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-02-23 1:00 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 4.06 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.06 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 1:00 PM		
		●	Message:Temperature Clear: 4.06 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 1:00 PM		
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.45 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 12:40 PM		
		●	Message:Temperature Alert: 1.45 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 12:40 PM		
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.66 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 12:17 PM		
		●	Message:Temperature Alert: 0.66 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 12:17 PM		
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.56 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 11:54 AM		
		●	Message:Temperature Alert: 1.56 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-23 11:54 AM		
2024-02-21 2:31 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.91 °C	8.0/2.0 °C	Alert Cleared
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.91 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-21 2:31 AM		
		●	Message:Temperature Clear: 2.91 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-21 2:31 AM		
		●	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.23 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-21 2:18 AM		
		●	Message:Temperature Alert: 1.23 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge	2024-02-21 2:18 AM		
2024-02-20 10:17 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	5.79 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-02-20 10:17 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 5.79 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-20 10:17 PM		Message:Temperature Clear: 5.79 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-20 10:08 PM		Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-02-20 10:08 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.42 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-20 10:08 PM		Message:Temperature Alert: 1.42 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

2024-02-20 7:34 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 3.94 °C 8.0/2.0 °C Alert Cleared

			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-02-20 7:34 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.94 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-20 7:34 PM		Message:Temperature Clear: 3.94 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-02-20 7:27 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.73 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-02-20 7:27 PM		Message:Temperature Alert: 1.73 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

2024-02-20 3:42 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.81 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-20 3:42 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.81 °C	8.0/1.0 °C	Alert Cleared
2024-02-20 2:42 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.93 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-15 6:08 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.73 °C	8.0/1.0 °C	Alert Cleared
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.73 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4 ● 2024-02-15 6:08 AM Message:Temperature Clear: 3.73 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 12.48 °C, still above the limit of 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4 ● 2024-02-15 5:50 AM Message:Temperature Alert: 12.48 °C, above the limit of 8.0 °C Alert Rule:Delp G040 +4
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 11.23 °C, still above the limit of 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4 ● 2024-02-15 5:29 AM Message:Temperature Alert: 11.23 °C, above the limit of 8.0 °C Alert Rule:Delp G040 +4
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 9.66 °C, above the limit of 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4 ● 2024-02-15 5:08 AM Message:Temperature Alert: 9.66 °C, above the limit of 8.0 °C Alert Rule:Delp G040 +4
2024-02-08 7:52 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.12 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-08 7:52 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.12 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-08 7:52 PM					Message:Temperature Clear: 2.12 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-08 7:39 PM					Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-08 7:39 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.65 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-08 7:39 PM					Message:Temperature Alert: 1.65 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-08 7:15 PM					Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-08 7:15 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.52 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-02-08 7:15 PM					Message:Temperature Alert: 1.52 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-02-02 9:22 AM Sudler 2012 CTSU Lab Fridge Temperature **Delp G040 -20** 5.37 °C -18.0/-30.0 °C

	2024-02-02 9:22 AM					Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:22 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 5.37 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 -20
	2024-02-02 9:22 AM					Message:Temperature Alert: 5.37 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20

	2024-02-02 9:01 AM					Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:01 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.41 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 -20
	2024-02-02 9:01 AM					Message:Temperature Alert: 4.41 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20

	2024-02-02 8:40 AM					Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:40 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.20 °C, above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 -20
	2024-02-02 8:40 AM					Message:Temperature Alert: 2.20 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20

2024-02-02 8:42 AM Sudler 2012 CTSU Lab Fridge Temperature **Temp out of Range -20 Freezer** 3.88 °C -18.0/-30.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:42 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.88 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Temp out of Range -20 Freezer
	2024-02-02 8:42 AM					Message:Temperature Alert: 3.88 °C, above the limit of -5.0 °C Alert Rule:Temp out of Range -20 Freezer
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:33 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.94 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Temp out of Range -20 Freezer
	2024-02-02 8:33 AM					Message:Temperature Alert: 3.94 °C, above the limit of -5.0 °C Alert Rule:Temp out of Range -20 Freezer
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:25 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 5.41 °C, above the limit of -5.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Temp out of Range -20 Freezer
	2024-02-02 8:25 AM					Message:Temperature Alert: 5.41 °C, above the limit of -5.0 °C Alert Rule:Temp out of Range -20 Freezer

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Lab Fridge
for
Sudler 2012 CTSU Lab Fridge

Mar 01, 2024 - Apr 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Lab Fridge" for Sudler 2012 CTSU Lab Fridge

Mar 01, 2024 - Apr 01, 2024

MACHINE DETAILS


Serial Number: **0081F90EC68E**
Location: **CTSU Rainbow Lab**
Description: **0081F90EC68E**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE











AVG: **4.21 °C** | MIN: **-0.18 °C** | MAX: **10.06 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-02-29 5:00 AM	4.98 °C
2024-03-01 5:00 AM	2.72 °C
2024-03-02 5:00 AM	4.84 °C
2024-03-03 5:00 AM	1.65 °C
2024-03-04 5:00 AM	3.44 °C
2024-03-05 5:00 AM	5.91 °C
2024-03-06 5:00 AM	5.07 °C
2024-03-07 5:00 AM	4.7 °C
2024-03-08 5:00 AM	4.19 °C
2024-03-09 5:00 AM	5.68 °C
2024-03-10 5:00 AM	1.96 °C
2024-03-11 5:00 AM	3.9 °C
2024-03-12 5:00 AM	3.67 °C
2024-03-13 5:00 AM	5.18 °C
2024-03-14 5:00 AM	5.91 °C
2024-03-15 5:00 AM	2.72 °C
2024-03-16 5:00 AM	3.92 °C
2024-03-17 5:00 AM	5.87 °C
2024-03-18 5:00 AM	2.1 °C
2024-03-19 5:00 AM	5.28 °C
2024-03-20 5:00 AM	1.86 °C
2024-03-21 5:00 AM	4.91 °C
2024-03-22 5:00 AM	4.37 °C
2024-03-23 5:00 AM	3.5 °C
2024-03-24 5:00 AM	4.62 °C
2024-03-25 5:00 AM	1.45 °C
2024-03-26 5:00 AM	5.09 °C
2024-03-27 5:00 AM	5.2 °C
2024-03-28 5:00 AM	1.96 °C
2024-03-29 5:00 AM	4.52 °C
2024-03-30 5:00 AM	5.55 °C
2024-03-31 5:00 AM	5.43 °C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-03-28 11:43 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	3.99 °C	8.0/2.0 °C	Alert Cleared
	2024-03-28 11:43 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.99 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 11:43 PM	Message:Temperature Clear: 3.99 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 11:30 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.05 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 11:30 PM	Message:Temperature Alert: 1.05 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
2024-03-28 4:42 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	3.9 °C	8.0/2.0 °C	Alert Cleared
	2024-03-28 4:42 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.90 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 4:42 PM	Message:Temperature Clear: 3.90 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 4:30 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.87 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 4:30 PM	Message:Temperature Alert: 0.87 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 4:08 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.85 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-28 4:08 PM	Message:Temperature Alert: 0.85 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
2024-03-28 4:20 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.61 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 4:20 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.61 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 3.61 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 4:13 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.95 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-28 2:09 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.74 °C 8.0/2.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 2:09 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.74 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 2.74 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 2:03 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.65 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.65 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 1:40 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.38 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.38 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-03-23 4:27 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.04 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-23 4:27 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.04 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-23 4:27 PM		Message:Temperature Clear: 2.04 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-23 4:20 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.59 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-23 4:20 PM		Message:Temperature Alert: 1.59 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

2024-03-22 6:45 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.59 °C 8.0/2.0 °C Alert Cleared

			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-22 6:45 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.59 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-22 6:45 AM		Message:Temperature Clear: 2.59 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-22 6:32 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.91 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-22 6:32 AM		Message:Temperature Alert: 1.91 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-22 6:09 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.59 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-22 6:09 AM		Message:Temperature Alert: 0.59 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

2024-03-22 5:57 AM Sudler 2012 CTSU Lab Fridge Temperature **Delp G040 +4** 1.98 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 5:57 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.98 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
	2024-03-22 5:57 AM					Message:Temperature Clear: 1.98 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 5:51 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.65 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
	2024-03-22 5:51 AM					Message:Temperature Alert: 0.65 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-21 6:59 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 3.97 °C 8.0/2.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 6:59 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.97 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-21 6:59 AM					Message:Temperature Clear: 3.97 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 6:52 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.98 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-21 6:52 AM					Message:Temperature Alert: 0.98 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-03-21 4:24 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.14 °C 8.0/2.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 4:24 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.14 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-21 4:24 AM					Message:Temperature Clear: 2.14 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 4:12 AM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.86 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-21 4:12 AM					Message:Temperature Alert: 1.86 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-03-20 6:16 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	4.02 °C	8.0/2.0 °C	Alert Cleared
	2024-03-20 6:16 PM	<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.02 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
	2024-03-20 6:16 PM	<p>Message:Temperature Clear: 4.02 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
	2024-03-20 5:50 PM	<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.84 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
	2024-03-20 5:50 PM	<p>Message:Temperature Alert: 1.84 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
	2024-03-20 5:27 PM	<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.48 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
	2024-03-20 5:27 PM	<p>Message:Temperature Alert: 1.48 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>				
2024-03-20 4:02 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.5 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 4:02 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.50 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
	2024-03-20 4:02 PM					Message:Temperature Clear: 3.50 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 3:55 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.63 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
	2024-03-20 3:55 PM					Message:Temperature Alert: 0.63 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-20 2:07 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 4.72 °C 8.0/2.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 2:07 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 4.72 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-20 2:07 PM					Message:Temperature Clear: 4.72 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 1:53 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.00 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-20 1:53 PM					Message:Temperature Alert: 1.00 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 1:30 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.98 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
	2024-03-20 1:30 PM					Message:Temperature Alert: 1.98 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-03-20 5:42 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.6 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.60 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge ● Message:Temperature Clear: 2.60 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.69 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge ● Message:Temperature Alert: 1.69 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
2024-03-20 4:01 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.98 °C	8.0/2.0 °C	Alert Cleared
						<ul style="list-style-type: none"> ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.98 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge ● Message:Temperature Clear: 2.98 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge ● Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.65 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge ● Message:Temperature Alert: 1.65 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
2024-03-20 2:00 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.25 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-19 4:24 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.08 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-19 4:24 PM Message:Temperature Clear: 2.08 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-19 4:12 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.91 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-19 4:12 PM Message:Temperature Alert: 0.91 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-03-19 12:32 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.78 °C	8.0/2.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-19 12:32 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.78 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-19 12:32 PM Message:Temperature Clear: 2.78 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-19 12:07 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.13 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-19 12:07 PM Message:Temperature Alert: 1.13 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-03-12 11:19 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.88 °C	8.0/2.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-12 11:19 AM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.88 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-12 11:19 AM Message:Temperature Clear: 2.88 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-03-12 11:04 AM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.73 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
						<p>● 2024-03-12 11:04 AM Message:Temperature Alert: 1.73 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

2024-03-03 10:41 PM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.18 °C 8.0/2.0 °C Alert Cleared

● 2024-03-03 10:41 PM Notification:Email, SMS, and App sent to Miranda Wright.
 Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.18 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
 Alert Rule:SUDLER 2012 CTSU lab Fridge

● 2024-03-03 10:41 PM Message:Temperature Clear: 2.18 °C, within the limit of 2.0 °C to 8.0 °C
 Alert Rule:SUDLER 2012 CTSU lab Fridge

● 2024-03-03 10:34 PM Notification:Email, SMS, and App sent to Miranda Wright.
 Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.47 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
 Alert Rule:SUDLER 2012 CTSU lab Fridge

● 2024-03-03 10:34 PM Message:Temperature Alert: 1.47 °C, below the limit of 2.0 °C
 Alert Rule:SUDLER 2012 CTSU lab Fridge

● 2024-03-03 10:09 PM Notification:Email, SMS, and App sent to Miranda Wright.
 Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.61 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
 Alert Rule:SUDLER 2012 CTSU lab Fridge

● 2024-03-03 10:09 PM Message:Temperature Alert: 1.61 °C, below the limit of 2.0 °C
 Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-03-03 10:49 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.19 °C 8.0/2.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-03 10:49 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.19 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Clear: 2.19 °C, within the limit of 2.0 °C to 8.0 °C			
	2024-03-03 10:49 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-03 10:23 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.82 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 1.82 °C, below the limit of 2.0 °C			
	2024-03-03 10:23 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-03 9:57 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.20 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 1.20 °C, below the limit of 2.0 °C			
	2024-03-03 9:57 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-03-03 6:37 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.29 °C	8.0/2.0 °C	Alert Cleared
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-03 6:37 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.29 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Clear: 2.29 °C, within the limit of 2.0 °C to 8.0 °C			
	2024-03-03 6:37 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-03 6:11 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.57 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 1.57 °C, below the limit of 2.0 °C			
	2024-03-03 6:11 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-03-03 4:35 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.52 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.52 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 2.52 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.63 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 0.63 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

2024-03-03 3:20 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 5.32 °C 8.0/2.0 °C Alert Cleared



						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 5.32 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 5.32 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge



						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.34 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 0.34 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge



2024-03-03 2:01 AM Sudler 2012 CTSU Lab Fridge Temperature **SUDLER 2012 CTSU lab Fridge** 2.52 °C 8.0/2.0 °C Alert Cleared



CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
		Notification:Email, SMS, and App sent to Miranda Wright.				
	2024-03-03 2:01 AM	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.52 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-03 2:01 AM	Message:Temperature Clear: 2.52 °C, within the limit of 2.0 °C to 8.0 °C				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
		Notification:Email, SMS, and App sent to Miranda Wright.				
	2024-03-03 1:54 AM	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.65 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-03 1:54 AM	Message:Temperature Alert: 0.65 °C, below the limit of 2.0 °C				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
		Notification:Email, SMS, and App sent to Miranda Wright.				
	2024-03-03 1:28 AM	Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.52 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
	2024-03-03 1:28 AM	Message:Temperature Alert: 0.52 °C, below the limit of 2.0 °C				
		Alert Rule:SUDLER 2012 CTSU lab Fridge				
2024-03-02 11:51 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.02 °C	8.0/2.0 °C	Alert Cleared





CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

	2024-03-02 11:51 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.02 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-02 11:51 PM		Message:Temperature Clear: 2.02 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

	2024-03-02 11:32 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.22 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-02 11:32 PM		Message:Temperature Alert: 1.22 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

	2024-03-02 11:07 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.95 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-02 11:07 PM		Message:Temperature Alert: 1.95 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

	2024-03-02 10:42 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.51 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
	2024-03-02 10:42 PM		Message:Temperature Alert: 1.51 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			

2024-03-02 10:03 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.31 °C	8.0/1.0 °C	Alert Cleared
	2024-03-02 10:03 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.31 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
	2024-03-02 10:03 PM		Message:Temperature Clear: 2.31 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			
	2024-03-02 9:56 PM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.38 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
	2024-03-02 9:56 PM		Message:Temperature Alert: 0.38 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			

2024-03-02 8:45 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.51 °C	8.0/2.0 °C	Alert Cleared
--------------------	-----------------------------	-------------	------------------------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

●	2024-03-02 8:45 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.51 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 8:45 PM	Message:Temperature Clear: 2.51 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				

●	2024-03-02 8:39 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.67 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 8:39 PM	Message:Temperature Alert: 1.67 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				

2024-03-02 6:12 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.28 °C	8.0/2.0 °C	Alert Cleared
●	2024-03-02 6:12 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.28 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 6:12 PM	Message:Temperature Clear: 2.28 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 5:58 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.61 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 5:58 PM	Message:Temperature Alert: 0.61 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 5:33 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.09 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 5:33 PM	Message:Temperature Alert: 1.09 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				

2024-03-02 3:32 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.12 °C	8.0/2.0 °C	Alert Cleared
--------------------	-----------------------------	-------------	------------------------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 3:32 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.12 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Clear: 2.12 °C, within the limit of 2.0 °C to 8.0 °C
	2024-03-02 3:32 PM					Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 1:53 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.91 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 1.91 °C, below the limit of 2.0 °C
	2024-03-02 1:53 PM					Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 1:28 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.87 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 0.87 °C, below the limit of 2.0 °C
	2024-03-02 1:28 PM					Alert Rule:SUDLER 2012 CTSU lab Fridge
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 1:01 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.00 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:SUDLER 2012 CTSU lab Fridge
						Message:Temperature Alert: 2.00 °C, below the limit of 2.0 °C
	2024-03-02 1:01 PM					Alert Rule:SUDLER 2012 CTSU lab Fridge
2024-03-02 1:34 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.9 °C	8.0/1.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 1:34 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.90 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.90 °C, within the limit of 1.0 °C to 8.0 °C
	2024-03-02 1:34 PM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-02 1:28 PM					Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.87 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.87 °C, below the limit of 1.0 °C
	2024-03-02 1:28 PM					Alert Rule:Delp G040 +4
2024-03-02 11:51 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.02 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
●	2024-03-02 11:51 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.02 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 11:51 AM		Message:Temperature Clear: 2.02 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 11:44 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.24 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 11:44 AM		Message:Temperature Alert: 0.24 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 11:19 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.77 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 11:19 AM		Message:Temperature Alert: 0.77 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 10:52 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.75 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge			
●	2024-03-02 10:52 AM		Message:Temperature Alert: 0.75 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-03-02 11:31 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.23 °C	8.0/1.0 °C	Alert Cleared
●	2024-03-02 11:31 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.23 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-02 11:31 AM		Message:Temperature Clear: 1.23 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			
●	2024-03-02 11:12 AM		Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.77 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-02 11:12 AM		Message:Temperature Alert: 0.77 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			
2024-03-02 10:39 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.18 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

●	2024-03-02 10:39 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.18 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
●	2024-03-02 10:39 AM	Message:Temperature Clear: 1.18 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4

●	2024-03-02 10:33 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.67 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4
●	2024-03-02 10:33 AM	Message:Temperature Alert: 0.67 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-02 9:29 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.04 °C	8.0/2.0 °C	Alert Cleared
●	2024-03-02 9:29 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.04 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 9:29 AM	Message:Temperature Clear: 2.04 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 9:22 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.98 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 9:22 AM	Message:Temperature Alert: 1.98 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				

2024-03-02 8:11 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.12 °C	8.0/2.0 °C	Alert Cleared
●	2024-03-02 8:11 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.12 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge				
●	2024-03-02 8:11 AM	Message:Temperature Clear: 2.12 °C, within the limit of 2.0 °C to 8.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge				

●	2024-03-02 7:58 AM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.12 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:SUDLER 2012 CTSU lab Fridge
●	2024-03-02 7:58 AM	Message:Temperature Alert: 0.12 °C, below the limit of 2.0 °C Alert Rule:SUDLER 2012 CTSU lab Fridge

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

2024-03-02 4:13 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.18 °C	8.0/2.0 °C	Alert Cleared
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 4:13 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.18 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Clear: 2.18 °C, within the limit of 2.0 °C to 8.0 °C			
	2024-03-02 4:13 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 12:31 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.30 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 0.30 °C, below the limit of 2.0 °C			
	2024-03-02 12:31 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 12:06 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.87 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 0.87 °C, below the limit of 2.0 °C			
	2024-03-02 12:06 AM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-01 11:41 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.18 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:SUDLER 2012 CTSU lab Fridge			
			Message:Temperature Alert: 0.18 °C, below the limit of 2.0 °C			
	2024-03-01 11:41 PM		Alert Rule:SUDLER 2012 CTSU lab Fridge			
2024-03-02 1:03 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.89 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 1:03 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.89 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:Delp G040 +4			
	2024-03-02 1:03 AM		Message:Temperature Clear: 1.89 °C, within the limit of 1.0 °C to 8.0 °C			
			Alert Rule:Delp G040 +4			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 12:50 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.62 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:Delp G040 +4			
	2024-03-02 12:50 AM		Message:Temperature Alert: 0.62 °C, below the limit of 1.0 °C			
			Alert Rule:Delp G040 +4			
2024-03-02 12:00 AM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.73 °C	8.0/1.0 °C	Alert Cleared
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-02 12:00 AM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.73 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:Delp G040 +4			
	2024-03-02 12:00 AM		Message:Temperature Clear: 1.73 °C, within the limit of 1.0 °C to 8.0 °C			
			Alert Rule:Delp G040 +4			
			Notification:Email, SMS, and App sent to Miranda Wright.			
	2024-03-01 11:54 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.23 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.			
			Alert Rule:Delp G040 +4			
	2024-03-01 11:54 PM		Message:Temperature Alert: 0.23 °C, below the limit of 1.0 °C			
			Alert Rule:Delp G040 +4			
2024-03-01 10:14 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.12 °C	8.0/2.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:Delp G040 +4</p> <p>Message:Temperature Clear: 2.08 °C, within the limit of 1.0 °C to 8.0 °C</p> <p>Alert Rule:Delp G040 +4</p> 						
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.46 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:Delp G040 +4</p> <p>Message:Temperature Alert: 0.46 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p> 						
2024-03-01 5:42 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	2.69 °C	8.0/2.0 °C	Alert Cleared
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.69 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>Message:Temperature Clear: 2.69 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> 						
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.05 °C, still below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>Message:Temperature Alert: 1.05 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> 						
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 9.09 °C, still above the limit of 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>Message:Temperature Alert: 9.09 °C, above the limit of 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> 						
<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.38 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>Message:Temperature Alert: 1.38 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> 						
2024-03-01 4:42 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	2.96 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 4:42 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 2.96 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 4:42 PM		Message:Temperature Clear: 2.96 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			

			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 4:28 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.04 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 4:28 PM		Message:Temperature Alert: 0.04 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			

2024-03-01 4:05 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	1.33 °C	8.0/1.0 °C	Alert Cleared
			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 4:05 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.33 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 4:05 PM		Message:Temperature Clear: 1.33 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			
			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 3:47 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 0.59 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 3:47 PM		Message:Temperature Alert: 0.59 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4			

2024-03-01 3:06 PM	Sudler 2012 CTSU Lab Fridge	Temperature	Delp G040 +4	3.69 °C	8.0/1.0 °C	Alert Cleared
			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 3:06 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.69 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 3:06 PM		Message:Temperature Clear: 3.69 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4			

			Notification:Email, SMS, and App sent to Miranda Wright.			
●	2024-03-01 2:44 PM		Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 9.09 °C, above the limit of 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab. Alert Rule:Delp G040 +4			
●	2024-03-01 2:44 PM		Message:Temperature Alert: 9.09 °C, above the limit of 8.0 °C Alert Rule:Delp G040 +4			

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-03-01 12:01 PM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	3.03 °C	8.0/2.0 °C	Alert Cleared
						<ul style="list-style-type: none"> <p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>2024-03-01 12:01 PM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 3.03 °C, within the limit of 2.0 °C to 8.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>2024-03-01 12:01 PM Message:Temperature Clear: 3.03 °C, within the limit of 2.0 °C to 8.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>2024-03-01 11:37 AM Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>2024-03-01 11:37 AM Message:Sudler 2012 CTSU Lab Fridge is reporting Temperature of 1.69 °C, below the limit of 2.0 °C. Sudler 2012 CTSU Lab Fridge is located in CTSU Rainbow Lab.</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p> <p>2024-03-01 11:37 AM Message:Temperature Alert: 1.69 °C, below the limit of 2.0 °C</p> <p>Alert Rule:SUDLER 2012 CTSU lab Fridge</p>
2024-03-01 5:25 AM	Sudler 2012 CTSU Lab Fridge	Temperature	SUDLER 2012 CTSU lab Fridge	4.3 °C	8.0/2.0 °C	Alert Cleared

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Lab Fridge
for
Sudler 2012 CTSU Lab Fridge

Apr 01, 2024 - May 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Lab Fridge" for Sudler 2012 CTSU Lab Fridge

Apr 01, 2024 - May 01, 2024

MACHINE DETAILS


Serial Number: **0081F90EC68E**
Location: **CTSU Rainbow Lab**
Description: **0081F90EC68E**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **4.35 °C** | MIN: **0.55 °C** | MAX: **9.21 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-03-31 5:00 AM	5.63°C
2024-04-01 5:00 AM	5.29°C
2024-04-02 5:00 AM	3.69°C
2024-04-03 5:00 AM	4.16°C
2024-04-04 5:00 AM	4.2°C
2024-04-05 5:00 AM	2.04°C
2024-04-06 5:00 AM	5.76°C
2024-04-07 5:00 AM	5.26°C
2024-04-08 5:00 AM	3.11°C
2024-04-09 5:00 AM	3.82°C
2024-04-10 5:00 AM	3.25°C
2024-04-11 5:00 AM	5.77°C
2024-04-12 5:00 AM	5.87°C
2024-04-13 5:00 AM	5.51°C
2024-04-14 5:00 AM	5.62°C
2024-04-15 5:00 AM	5.84°C
2024-04-16 5:00 AM	4.14°C
2024-04-17 5:00 AM	3.88°C
2024-04-18 5:00 AM	5.16°C
2024-04-19 5:00 AM	4.77°C
2024-04-20 5:00 AM	5.04°C
2024-04-21 5:00 AM	5.6°C
2024-04-22 5:00 AM	4.8°C
2024-04-23 5:00 AM	3.63°C
2024-04-24 5:00 AM	3.44°C
2024-04-25 5:00 AM	2.97°C
2024-04-26 5:00 AM	4.16°C
2024-04-27 5:00 AM	5.18°C
2024-04-28 5:00 AM	4.79°C
2024-04-29 5:00 AM	3.17°C
2024-04-30 5:00 AM	5.79°C

ALERTS

No Alert Events for selected date range.