

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -65°C to -90°C

111768301191108 (-80 Freezer Delp)

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 5 May Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Thu	5/1/2025	1124	-79	-80	MW	
Fri	5/2/2025	0810	-79	-79	MW	
Mon	5/5/2025	0805	-79	-79	MW	
Tue	5/6/2025	0815	-79	-79	MW	
Wed	5/7/2025	0803	-79	-79	MW	
Thu	5/8/2025	0810	-79	-80	MW	
Fri	5/9/2025	0838	-79	-80	MW	
Mon	5/12/2025	1000	-79	-79	ADK	
Tue	5/13/2025	0813	-79	-80	MW	
Wed	5/14/2025	0909	-80	-80	MW	
Thu	5/15/2025	0744	-79	-80	MW	
Fri	5/16/2025	0833	-79	-80	MW	
Mon	5/19/2025	0733	-79	-80	MW	
Tue	5/20/2025	0800	-79	-80	MW	
Wed	5/21/2025	0818	-79	-79	MW	
Thu	5/22/2025	0830	-79	-74	ADK	
Fri	5/23/2025	0830	-80	-79	ADK	
Mon	5/26/2025					Holiday ADK
Tue	5/27/2025	0800	-79	—	ADK	display not working ADK
Wed	5/28/2025	0920	-79	—	ADK	display not working ADK
Thu	5/29/2025	0815	-79	—	ADK	display not working ADK
Fri	5/30/2025	0745	-79	—	ADK	display not working ADK
Mon	6/2/2025					

Key: * Heavy usage