

Thermometer Performance Report

Customer: University of KS Medical Center-OTSU Fairway		Contract:		Shelby Williams				
Street: 4350 Shawnee Mission Pkwy		Phone #		(913) 588-0980				
City/ST/Zip: Fairway, KS 66205		±: 0.3-C-Pass		P= Pass F=Fail				
Asset#	Model#	MFG	SIN	Temperature (Celsius)		Date Performed	Next Due	Pass/Fail
				Unit Temp	Certified Temp			
WA-001	901053	Weich Allyn	(21)20467939	35.4°	35.5°	07/03/2023	7/2024	P
WA-002	901053	Weich Allyn	(21)20258174	34.6°	34.6°	07/03/2023	7/2024	P
WA-003	901053	Weich Allyn	(21)20258157	33.3°	33.5°	07/03/2023	7/2024	P
WA-004	901053	Weich Allyn	(21)21075567	37.1°	37.3°	07/03/2023	7/2024	P
WA-005	901053	Weich Allyn	(21)20258164	36.5°	36.7°	07/03/2023	7/2024	P
WA-006	901053	Weich Allyn	(21)21055055	35.4°	35.2°	07/03/2023	7/2024	P
WA-007	901053	Weich Allyn	(21)20258280	35.0°	35.0°	07/03/2023	7/2024	P

FLUKE-521I Thermometer SN: 15100368 EX: 02/01/2024

Technician: Andrew Ravitch

Notes: Next Due Date: 07/2024

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa, Lenexa KS, 66212 913-403-0088