

**COBB Refrigeration & Laboratory Services LLC**

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

**CERTIFICATION OF ACCURACY**

**Customer Information**

**Company Name: University of KS Medical Center-CTSU**

**Company Address: 3901 Rainbow Blvd Kansas City, KS 66103**

**Contact Person: Shelby Williams**

**Phone Number: 913-574-3018**

---

**Thermometer Information:** Built-In

**Manufacturer:** Fisher Scientific

**Model:** TSX60086A

**Serial No:** 111768301191108

**Department/Location:** CTSU Rainbow

**Asset #** N/A

**Set Point:** -80°C

**Certified Temperature:** -79.1°C

**Observed Temperature:** -79°C

**Elemental Monitor:** N/A

**Observed Accuracy:** .1°C

**Passed/Failed:** Pass

**Certification Date:** 1/10/2024

**Certification Due:** 1/10/2025

---

**Certifying Equipment**

**Manufacturer:** Fluke

**Model:** 54iiB

**Serial No:** 86470112

**Due Date:** 04/18/2024

---

**Service Technician:** Nicholas Brown  
**(Print)**

**Date:** 1/10/2024

**Service Technician:** \_\_\_\_\_  
**(Sign)**

