

**Thermometer Performance Report**

Customer:		University of Kansas Health Syst-CTSU (Fairway)								Contact:		Shelby Williams			
Street:		4350 Shawnee Mission Pkw								Phone #		913-574-3018			
City/ST/Zip:		Fairway, KS 66205								+/- 2 C=Pass		P= Pass		F=Fail	
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)				
											Enviro Alert	Unit Temp	Certified Temp	Date Performed	Date Due
F-1	989	THERMO	412269-86	P	P	P	P	P	P	P	-82	-79	-80.1	1/25/2023	1/2024
F-3	989	THERMO	841803-1140	P	P	P	P	P	P	P	-80	-78	-80.0	1/25/2023	1/2024
F-4	ABT-HC-10PG	ABS	ABS-951703280749PW-G-1801	P	P	P	P	P	P	P	3	5.4	5.1	1/25/2023	1/2024
F-5	2-UCF-4	AEGIS	7075159	P	P	P	P	P	P	P	-20	-17.9	-18.5	1/25/2023	1/2024
F-13	FRT21FG3DW5	FRIGIDAIRE	LA44310002	P	P	P	P	P	P	P	-17	-16.8	-17.4	1/25/2023	1/2024
F-14	FRT21FG3DW5	FRIGIDAIRE	LA44310002	P	P	P	P	P	P	P	3	4.5	4.7	1/25/2023	1/2024
F-3300	ROOM TEMP	ENVIRO ALERT	ROOM 3300	P	P	P	P	P	P	P	20.7	20.1	21.2	1/25/2023	1/2024
F-3537	ROOM TEMP	ENVIRO ALERT	ROOM 3537	P	P	P	P	P	P	P	20.5	20.5	22.2	1/25/2023	1/2024
F-15	133000	BOEKEL(INCUBATOR)	192423688	P	P	P	P	P	P	P	NA	37.1	37.2	1/25/2023	1/2024

Technician:

Test equipment used

Notes:

**Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088**