

Service Report



WORK ORDER ID: CWKD6975511

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIVERSITY OF KS MED CTR-CTSU | **DEPARTMENT:** 146148-UNIVERSITY OF KS MED CTR-CTSU-BIOMED | **ACCOUNT NUMBER:** 146148

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Mail Stop 6011 Kansas City, KS 66103

SHIPPING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

AGILITY ASSET ID : 1887913 | **SERIAL #:** 192423688 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: INCUBATORS, LABORATORY, AEROBIC | **MANUFACTURER:** BOEKEL SCIENTIFIC INC | **MODEL:** DIGITAL INCUBATOR

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-22-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.4	M-F 8-5	01/22/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
DM1	0286	DIGITAL MULTIMETER	04/04/24

CHECKLIST

PROCEDURE: L1200 L1

QUESTION

Alarm - High Temperature
Alarm - Low Temperature
CES THERM.
Clinical- Audio
Clinical- Visual
CUST. THERM.
DESIRED TEMP
Electrical Safety Inspection
Mechanical Safety Inspection
Operational Check
Patient- Audio
Patient- Visual
Remote
Safety- Audio
Safety- Visual

ANSWER

Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass

Customer Signature:

Stephanie Dye

Technician Signature:

Date:

1/22/25

Date: