

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Medication Room Refrigerator

Refrigerator or Freezer ID/Location:
CTSU FAIRWAY

Month 8 August Year 2025 Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Fri	8/1/2025	0853	6.14	3.19	7.90	BEM	
Mon	8/4/2025	0656	6.34	3.21	8.44	Yds	
Tue	8/5/2025	0640	4.76	3.21	8.48	Yds	* - 4 AM
Wed	8/6/2025	0900	7.23	3.22	8.48	BEM	
Thu	8/7/2025	0715	6.91	3.07	8.66	Yds	* (8.56°C at 2219)
Fri	8/8/2025						
Mon	8/11/2025	0640	5.48	3.03	7.77	Yds	
Tue	8/12/2025	0917	7.23	3.06	8.27	Yds	* 9:57=8.21°C High use?
Wed	8/13/2025	0700	3.41	3.07	8.27	Yds	
Thu	8/14/2025	0728	7.23	3.01	8.24	Yds	* 5:48:32=8.29°C
Fri	8/15/2025	0848	6.98	2.99	8.29	BEM	
Mon	8/18/2025	0717	6.02	2.97	8.60	Yds	Temp 8.53°C at 1042:36 PM EDT
Tue	8/19/2025	0646	6.91	2.97	7.45	Yds	
Wed	8/20/2025	0715	5.41	3.03	7.47	Yds	
Thu	8/21/2025	0656	5.07	3.03 2.99	8.60	Yds	*
Fri	8/22/2025						
Mon	8/25/2025	0650	4.47	3.34	7.56	Yds	
Tue	8/26/2025	0651	3.51	2.97	7.54	Yds	
Wed	8/27/2025	0720	4.82 6.91	2.97	8.54	Yds	* 02:53:04 8/27 OOR
Thu	8/28/2025	0700	4.82 8.47	2.96	8.54	Yd	Temp 8.47°C @ 02:53:05
Fri	8/29/2025	0858	3.22	2.98	8.29	BEM	
Mon	9/1/2025						
Tue	9/2/2025						See Sept. 2025

Key: * Heavy usage