

Service Report



WORK ORDER ID: CWKD6948033

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 2012 Suder Kansas City, KS 66103

SHIPPING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITI ASSET ID : 1887947 | **SERIAL #:** 9303139 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**
ASSET TYPE: INCUBATORS, LABORATORY, AEROBIC | **MANUFACTURER:** QUINCY LAB | **MODEL:** 10-140

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-17-2025

LABOR

| SERVICE PROVIDER | CATEGORY | HOURS | PERIOD DEFINITION | SERVICE DATE |
|------------------|----------------|-------|-------------------|--------------|
| Stephanie Dye | PM Not Covered | 0.5 | M-F 8-5 | 01/16/25 |

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

| PREFIX | UNIT | DESCRIPTION | CALIBRATION DATE |
|--------|------|--------------------|------------------|
| DM1 | 0286 | DIGITAL MULTIMETER | 04/04/24 |

CHECKLIST

PROCEDURE: L1200 L1

QUESTION

Alarm - High Temperature
Alarm - Low Temperature
CES THERM.
Clinical- Audio
Clinical- Visual
CUST. THERM.
DESIRED TEMP
Electrical Safety Inspection
Mechanical Safety Inspection
Operational Check
Patient- Audio
Patient- Visual
Remote
Safety- Audio
Safety- Visual

ANSWER

Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass
Pass

Customer Signature:

Stephanie Dye

Technician Signature:

Date:

4th Febuary 2025

Date:

Service Report



WORK ORDER ID: CWKD6948021

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 2012 Suder Kansas City, KS 66103

SHIPPING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITI ASSET ID : 1887948 | **SERIAL #:** 192423685 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: INCUBATORS, LABORATORY, THERMOCYCLING | **MANUFACTURER:** BOEKEL SCIENTIFIC INC | **MODEL:** J1672-1

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-17-2025

LABOR

| SERVICE PROVIDER | CATEGORY | HOURS | PERIOD DEFINITION | SERVICE DATE |
|------------------|----------------|-------|-------------------|--------------|
| Stephanie Dye | PM Not Covered | 0.5 | M-F 8-5 | 01/16/25 |

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

| PREFIX | UNIT | DESCRIPTION | CALIBRATION DATE |
|--------|------|-------------------------|------------------|
| SB7 | 0102 | SAFETY ANALYZER, ESA620 | 07/24/24 |
| DM1 | 0286 | DIGITAL MULTIMETER | 04/04/24 |

CHECKLIST

PROCEDURE: L1200 L1

QUESTION

ANSWER

Alarm - High Temperature

Pass

Alarm - Low Temperature

Pass

CES THERM.

Pass

Clinical- Audio

Pass

Clinical- Visual

Pass

CUST. THERM.

Pass

DESIRED TEMP

Pass

Electrical Safety Inspection

Pass

Mechanical Safety Inspection

Pass

Operational Check

Pass

Patient- Audio

Pass

Patient- Visual

Pass

Remote

Pass

Safety- Audio

Pass

Safety- Visual

Pass

Customer Signature:

Date:

Stephanie Dye

4th February 2025

Technician Signature:

Date: