

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month **6** June Year **2025** Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon	6/2/2025	1024	-23.9	-27.6	-22.1	MW	
Tue	6/3/2025	0830	-23.6	-27.4	-22.4	ADK	
Wed	6/4/2025	0754	-25.4	-27.4	-22.4	MW	
Thu	6/5/2025	0807	-24.2	-27.1	-22.3	MW	
Fri	6/6/2025	0727	-24.2	-28.0	-11.3	MW	
Mon	6/9/2025	1000	-25.0	-29.0	-16.5	ADK	
Tue	6/10/2025	0747	-25.4	-28.9	-23.2	MW	
Wed	6/11/2025	0800	-23.3	-28.9	-23.3	ADK	
Thu	6/12/2025	0908	-27.4	-28.8	-22.9	MW	
Fri	6/13/2025	0830	-24.1	-28.9	-23.3	ADK	
Mon	6/16/2025	0826	-24.2	-29.0	-23.4	MW	
Tue	6/17/2025	0800	-25.1	-29.0	-23.5	ADK	
Wed	6/18/2025	0918	-24.1	-29.0	-22.3	MW	
Thu	6/19/2025						Closed for Holiday MW
Fri	6/20/2025	0930	-26.8	-29.0	-18.1	MW	
Mon	6/23/2025	0836	-26.0	-28.3	-23.8	MW	
Tue	6/24/2025	0805	-24.0	-28.1	-23.9	MW	
Wed	6/25/2025	0840	-26.6	-28.2	-23.9	MW	
Thu	6/26/2025	0705	-24.1	-29.1	-17.3	MW	
Fri	6/27/2025	0831	-24.7	-29.5	-23.4	MW	
Mon	6/30/2025	1040	-25.1	-29.5	-23.3	MW	
Tue	7/1/2025					N/A	
Wed	7/2/2025					N/A	

Key: * Heavy usage