

CERTIFICATION OF ACCURACY

Customer Information

Company Name: UKMC- Swope CTSU

Company Address: 4001 Dr. Martin Luther King Jr. Blvd, Kansas City, MO, 64130

Contact Person: Autumn Beauchamp abeauchamp@kumc.edu

Manufacturer: Fisher Brand

Model: GTFBA05COSA

Serial No: 300422763

Asset: 1887895

Set Point (°C): N/A

Certified Temperature (°C): 1.5

Observed Temperature (°C): N/A

Observed Accuracy (°C): N/A

Passed/Failed: Passed

Certification Date: 01/12/2026

Certification Due: 01/2027

Certifying Equipment

Manufacturer: Fluke

Model: 52II

Serial No: 66620177MV

Due Date: 04/21/2026

Service Technician: Robert McElwee
(Print)

Date: 01/12/2026

Service Technician: 
(Sign)