

Service Report



WORK ORDER ID: CWKD4080830

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy. Ste. #1360 Fairway, KS 66205

AGILITI ASSET ID : 1887924 | SERIAL #: 002535 | VENDOR SITE ID: | CUSTOMER ASSET ID:

ASSET TYPE:ASPIRATORS,AIRWAY | MANUFACTURER: PRECISION MEDICAL INC | MODEL: POWERVAC PM61

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887924

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Marcus Boatwright	PM Not Covered	0.35	M-F 8-5	01/23/24

SERVICE PROVIDED DETAILS

Completed SM passed all inspection testing 156/20 ES

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
DP1	252	DPM	05/19/23
SB6	223	SAFETY ANALYZER, FLUKE ESA612	04/25/23

CHECKLIST

PROCEDURE: SUC01 L1

QUESTION	ANSWER
CHNG FILTERS	NA
CONDITION	Pass
Electrical Safety Inspection	Pass
GAUGES	Pass
LUBE MOTOR/PUMP	Yes
MAX VACUUM	-63
Mechanical Safety Inspection	Pass
Operational Check	Pass
OVFLOW PROT	Pass
RISE RATE	N/A

Customer Signature:

Date:

Technician Signature:

Date: