

Sudler 2012 CTSU Lab Fridge
for
Sudler 2012 CTSU Lab Fridge

May 01, 2024 - Jun 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Lab Fridge" for Sudler 2012 CTSU Lab Fridge

May 01, 2024 - Jun 01, 2024

MACHINE DETAILS


Serial Number: **0081F90EC68E**
Location: **CTSU Rainbow Lab**
Description: **0081F90EC68E**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **4.38 °C** | MIN: **0.67 °C** | MAX: **8.15 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-05-01 12:00 AM	4.23°C
2024-05-02 12:00 AM	5.55°C
2024-05-03 12:00 AM	4.62°C
2024-05-04 12:00 AM	5.23°C
2024-05-05 12:00 AM	2.96°C
2024-05-06 12:00 AM	2.93°C
2024-05-07 12:00 AM	3.69°C
2024-05-08 12:00 AM	2.77°C
2024-05-09 12:00 AM	3.52°C
2024-05-10 12:00 AM	3.26°C
2024-05-11 12:00 AM	3.6°C
2024-05-12 12:00 AM	3.21°C
2024-05-13 12:00 AM	2.56°C
2024-05-14 12:00 AM	3.38°C
2024-05-15 12:00 AM	3.6°C
2024-05-16 12:00 AM	5.58°C
2024-05-17 12:00 AM	5.18°C
2024-05-18 12:00 AM	3.86°C
2024-05-19 12:00 AM	5.55°C
2024-05-20 12:00 AM	4.48°C
2024-05-21 12:00 AM	2.59°C
2024-05-22 12:00 AM	4.35°C
2024-05-23 12:00 AM	5.88°C
2024-05-24 12:00 AM	5.43°C
2024-05-25 12:00 AM	3.97°C
2024-05-26 12:00 AM	5.59°C
2024-05-27 12:00 AM	5.26°C
2024-05-28 12:00 AM	4.95°C
2024-05-29 12:00 AM	3.52°C
2024-05-30 12:00 AM	4.16°C
2024-05-31 12:00 AM	3.59°C

ALERTS

No Alert Events for selected date range.