

CERTIFICATION OF ACCURACY

Customer Information

Company Name: KUMC Swope CTSU

Company Address: 4001 Dr. Martin Luther King Jr. Blvd, Kansas City, MO 64130

Contact Person/ Phone Number: Autumn Beauchamp abeauchamp@kumc.edu

Manufacturer: Hillrom

Model: Spot Vital Signs 4400

Serial No: 100087194922

SPO2% Set Values: 60bpm 97%O2

SPO2% Read Values: 60bpm 97%O2

NIBP Set Values: 120/80

NIBP Read Values: 120/83

Temperature Meter Value: 36.7C

Temperature Read Value: 35.5C

Passed/Failed: Pass

Certification Date: 1/12/2026

Certification Due: 1/2027

Certifying Equipment

Manufacturer: Fluke

Model: Prosim8

Serial No: 3990033

Service Technician: Derek Hulett
(Print)

Date: 1/12/2026

Service Technician: Derek Hulett
(Sign)