

**Blood Pressure Monitor Performance Report**

Customer:	KU-CTSU Fairway	Contact:	Shelby Willams
Street:	4350 Shawnee Mission Parkway 3rd Floor	Phone #	913-574-3018
City/ST/Zip:	Fairway, KS 66205	P= Pass F=Fail	

MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	97%	
Hillrom	(91)73WX	(21)100093400821	P	P	P	P	P	P	122/82	60	202/149	130	60/32	40	97%	1/26/2023
Hillrom	(91)73WX	(21)100093500821	P	P	P	P	P	P	120/80	60	120/80	130	60/32	40	97%	1/26/2023
Hillrom	(91)73WX	(21)100093440821	P	P	P	P	P	P	121/82	60	121/82	130	61/32	40	97%	1/26/2023
Hillrom	(91)73WX	(21)100093430821	P	P	P	P	P	P	121/82	60	202/149	130	62/32	40	97%	1/26/2023
Hillrom	(91)73WX	(21)100093510821	P	P	P	P	P	P	121/81	60	201/149	130	62/32	40	97%	2/6/2023
Hillrom	(91)73WX	(21)100093460821	P	P	P	P	P	P	<b>OUT OF SERVICE</b>							
Hillrom	(91)73WX	(21)100093420821	P	P	P	P	P	P	122/80	60	201/149	130	60/32	40	97%	2/6/2023
Hillrom	(91)73WX	(21)100093450821	P	P	P	P	P	P	121/81	60	202/151	130	60/32	40	97%	1/26/2023

Technicians: **JC**  
 Initials:

Technician: Jim Cowart  
**Date: 01/26/2023**

Test equipment used  
**FLUKE- ProSim 8 S#399003 Expires:09/07/2023**

Notes: Next Due Date: 01/2024  
**Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088**

**Blood Pressure Monitor Performance Report**

Customer:		KU-CTSU Fairway			Contact:		Shelby Willams									
Street:		4350 Shawnee Mission Parkway 3rd Floor			Phone #		913-574-3018									
City/ST/Zip:		Fairway, KS 66205			P= Pass F=Fail											
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	97%	

**Blood Pressure Monitor Performance Report**

Customer:		KU-CTSU Fairway			Contact:		Shelby Williams									
Street:		4350 Shawnee Mission Parkway 3rd Floor			Phone #		913-574-3018									
City/ST/Zip:		Fairway, KS 66205			P= Pass F=Fail											
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	97%	