COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Scale Performance Report

Customer Information

Company Name: KU Med Center

Company Address: 3901 Rainbow, Kansas City MO

Contact Person/ Phone Number: Shelby Williams, 913-574-3018,

swilliams46@kumc.edu

(Sign)

Manufacturer: Health-o-Meter		Model: 500KL		Serial No: 500 0178353	
Physical Condition:					
Chassis/ Housing:		Pass	Х	Fail	
Controls & Switches:		Pass	Х	Fail	
Power Cord & Strain Relief:		Pass	Х	Fail	
Hardware Integrity:		Pass	X	Fail	
Indicators & Display Operation:		Pass	Х	Fail	
Mfg. Labeling Intact:		Pass	Х	Fail	
Operation:		Pass	Х	Fail	
Operation:					
Measured Values	Set:	50lbs	Measured: 50.0lbs		
	Set:	250lbs	Meas	sured: 250.4lbs	
	Set:	500lbs	Meas	ured: 500.2ibs	
Certification Date: 1/11/2024				Certification Due: 1/2025	
21 20 20 20 10 1					
Certifying Equipment					
Kansas Metrology Laboratory W	eight Set Cert	ificate Num	ber: K1	.6924 Date Due: 4/3/24	
Method: Certified per manufact					
Service Technician: Nick Brown				Date: 1/11/24	-31-
Service Technician:	~		_		