

Service Report



WORK ORDER ID: CWKD7103145

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 2012 Suder Kansas City, KS 66103

SHIPPING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITY ASSET ID : 1887877 | **SERIAL #:** 12223000333 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: ELECTROCARDIOGRAPHS, MULTICHANNEL, INTERPRETIVE | **MANUFACTURER:** WELCH ALLYN INC | **MODEL:** ELI 280

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-17-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.4	M-F 8-5	01/16/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
PR1	0105	SIMCUBE	12/04/24
SA3	0114	SAFETY ANALYZER	03/20/24

CHECKLIST**PROCEDURE: M1230 L1****QUESTION****ANSWER**

CABLES	Pass
CALIBRATION	Pass
CLEAN	Yes
CLEAN	Yes
CLEAN/LUBE REC.	Yes
Clinical- Audio	Pass
Clinical- Visual	Pass
CONNECTORS	Pass
Electrical Safety Inspection	Pass
EXT CONDITION	Pass
MECH	Pass
Mechanical Safety Inspection	Pass
Operational Check	Pass
Patient- Audio	Pass
Patient- Visual	Pass
PLUGS	Pass
Remote	Pass
Safety- Audio	Pass
Safety- Visual	Pass
SPEED	Pass
STYLUS/PRINTHEA	Yes
WAVEFORM	Pass

Customer Signature:

Date:

Stephanie Dye

1-16-2025

Technician Signature:

Date:

Service Report



WORK ORDER ID: CWKD6945870

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 2012 Suder Kansas City, KS 66103

SHIPPING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITY ASSET ID : 1887878 | **SERIAL #:** 11922000794 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: ELECTROCARDIOGRAPHS, MULTICHANNEL, INTERPRETIVE | **MANUFACTURER:** WELCH ALLYN INC | **MODEL:** ELI 280

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-17-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.4	M-F 8-5	01/17/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME.

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
PR1	0105	SIMCUBE	12/04/24

CHECKLIST

PROCEDURE: M1230 L1

QUESTION

ANSWER

CABLES

Pass

CALIBRATION

Pass

CLEAN

Yes

CLEAN

Yes

CLEAN/LUBE REC.

Yes

Clinical- Audio

Pass

Clinical- Visual

Pass

CONNECTORS

Pass

Electrical Safety Inspection

Pass

EXT CONDITION

Pass

MECH

Pass

Mechanical Safety Inspection

Pass

Operational Check

Pass

Patient- Audio

Pass

Patient- Visual

Pass

PLUGS

Pass

Remote

Pass

Safety- Audio

Pass

Safety- Visual

Pass

SPEED

Pass

STYLUS/PRINTHEA

Yes

WAVEFORM

Pass

Customer Signature:

Date:

Stephanie Dye

4th Febuary 2025

Technician Signature:

Date: