

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month

7

July

Year

2025

Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Tue	7/1/2025	1244	-22.5	-29.4	-22.5	MW	
Wed	7/2/2025	0723	-23.6	-27.5	-15.7	MW	
Thu	7/3/2025	0717	-23.8	-28.5	-10.2	MW	
Fri	7/4/2025					MW	Holiday
Mon	7/7/2025	0750	-27.6	-28.7	-17.6	ADK	
Tue	7/8/2025	0719	-25.7	-28.7	-25.7	MW	
Wed	7/9/2025	0728	-25.1	-28.3	-22.8	MW	
Thu	7/10/2025	0930	-25.5	-28.4	-23.4	MW	
Fri	7/11/2025	1150	-26.0	-28.5	-23.5	MW	
Mon	7/14/2025	0745	-23.7	-28.4	-23.5	ADK	
Tue	7/15/2025	0745	-24.2	-28.6	-22.3	MW	
Wed	7/16/2025	0735	-25.6	-28.2	-21.3	MW	
Thu	7/17/2025	0800	-25.2	-28.3	-23.3	ADK	
Fri	7/18/2025	0737	-27.1	-28.2	-23.3	MW	
Mon	7/21/2025	0800	-27.7	-28.2	-23.2	MW	
Tue	7/22/2025	0737	-24.6	-28.3	-23.4	MW	
Wed	7/23/2025	0730	-24.4	-28.2	-23.3	ADK	
Thu	7/24/2025	0722	-25.8	-28.2	-23.5	MW	
Fri	7/25/2025	1003	-25.0	-28.3	-22.6	MW	
Mon	7/28/2025	0850	-23.4	-28.6	-17.8	ADK	
Tue	7/29/2025	0808	-25.0	-28.3	-22.6	MW	
Wed	7/30/2025	0754	-25.3	-27.9	-22.5	MW	
Thu	7/31/2025	0736	-27.3	-28.2	-22.4	MW	

Key: * Heavy usage