

# COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

## Scale Performance Report

### Customer Information

**Company Name:** KU Med Center

**Company Address:** 3901 Rainbow, Kansas City MO

**Contact Person/ Phone Number:** Shelby Williams, 913-574-3018,  
swilliams46@kumc.edu

---

**Manufacturer:** Health-o-Meter

**Model:** 500KL

**Serial No:** 500 0186170

### Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Operation:	Pass	X	Fail

### Operation:

Measured Values	Set:	50lbs	Measured:	49.8lbs
	Set:	250lbs	Measured:	247.0lbs
	Set:	500lbs	Measured:	493.0lbs

**Certification Date:** 1/11/2024

**Certification Due:** 1/2025

---

### Certifying Equipment

Kansas Metrology Laboratory Weight Set Certificate Number: K16924 Date Due: 4/3/24

Method: Certified per manufacturer's specifications and procedures traceable to NIST

---

**Service Technician:** Nick Brown

**Date:** 1/11/24

**Service Technician:**  
(Sign)

