

Centrifuge Performance Report

Company Name: UKMC- Swope CTSU

Company Address: 4001 Dr. Martin Luther King Jr. Blvd, Kansas City, MO, 64130

Contact Person: Autumn Beauchamp abeauchamp@kumc.edu

Equipment Information:

Manufacturer: Beckman Coulter Model: X-30R Serial No: ALZ19F009 Asset: 1887892

Physical Condition:

Chassis/ Housing:	Pass X	Fail
Controls & Switches:	Pass X	Fail
Power Cord & Strain Relief:	Pass X	Fail
Hardware Integrity:	Pass X	Fail
Indicators & Display Operation:	Pass X	Fail
Mfg. Labeling Intact:	Pass X	Fail

Operation:

Timer Accuracy	Pass: X	Fail:
Lid Latch Safety Interlock	Pass: X	Fail:
Brake	Pass: X	Fail:
Speed Check Low (rpm)	Set: 1000	Measured: 1002
Speed Check Med (rpm)	Set: 2000	Measured: 2000
Speed Check High (rpm)	Set: 4000	Measured: 4001
Temperature (C)	Set: 8	Measured: 7.5

Certification Date: 01/12/2026

Certification Due: 01/2027

Certifying Equipment

Manufacturer: Fluke Date: 04/21/2026

Manufacturer: Traceable Date: 05/08/2027

Manufacturer: Monarch Date: 06/18/2026

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Robert McElwee

Date: 01/12/2026

Service Technician: 