

## Scale Performance Report

### Customer Information

**Company Name:** KUMC Swope CTSU

**Company Address:** 4001 Dr. Martin Luther King Jr. Blvd, Kansas City, MO 64130

**Contact Person/ Phone Number:** Autumn Beauchamp [abeachuamp@kumc.edu](mailto:abeachuamp@kumc.edu)

---

**Manufacturer:** Health-O-Meter

**Serial No:** 5002621203362

### Physical Condition:

Chassis/ Housing:	Pass	<b>X</b>	Fail
Controls & Switches:	Pass	<b>X</b>	Fail
Power Cord & Strain Relief:	Pass	<b>X</b>	Fail
Hardware Integrity:	Pass	<b>X</b>	Fail
Indicators & Display Operation:	Pass	<b>X</b>	Fail
Mfg. Labeling Intact:	Pass	<b>X</b>	Fail
Operation:	Pass	<b>X</b>	Fail

### Operation:

Measured Values	Set:	50lb	Measured: 50.2lb
	Set:	250lb	Measured: 250.6lb
	Set:	500lb	Measured: 500.6lb

**Certification Date:** 1/12/2026

**Certification Due:** 1/2027

---

### Certifying Equipment

Accurate Superior Scale 50lb weights SN 26, 28, 29, 30, 31, 32, 33, 35, 36, 37 Due 2/21/2026

**Method:** Certified per manufacturer's specifications and procedures traceable to NIST

---

**Service Technician:** Derek Hulett

**Date:** 1/12/2026

**Service Technician:** Derek Hulett  
(Sign)