

**COBB Refrigeration & Laboratory Services LLC**  
8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

## CERTIFICATION OF ACCURACY

### Customer Information

**Company Name:** KU Med Center CTSU

**Company Address:** 3901 Rainbow, Kansas City, KS

**Contact Person/ Phone Number:** Danielle Hoage 913-574-3059

---

### Thermometer Information

**Manufacturer:** Covidien

**Model:** Filac EZ

**Serial No:** SMF0812-001

**Department:** CTSU

	High	Medium	Low
<b>Certified Temperature (°C):</b>	41.5	38.9	36.9
<b>Observed Temperature (°C):</b>	41.3	38.9	36.4
<b>Observed Accuracy (°C):</b>	0.2	0	0.5

**Passed/Failed:** Pass

**Certification Date:** 1/10/2024

**Certification Due:** 1/2025

---

### Certifying Equipment

**Manufacturer:** Fluke

**Model:** 52II

**Serial No:** 15100368

**Due Date:** 2/1/2024

---

**Service Technician:** Derek Hulett  
(Print)

**Date:** 1/10/2024

**Service Technician:**  
(Sign)

