

Report of Calibration
GLOBAL CALIBRATION LABORATORY at
FLUKE BIOMEDICAL
 2 Science Road
 Glenwood, IL 60425-1586 USA

CUSTOMER ASSET NUMBER: N/A	CUSTOMER INFORMATION COBB REFRIGERATION LLC LENEXA, KS
MANUFACTURER: FLUKE BIOMEDICAL	
MODEL NUMBER: PROSIM 8	
SERIAL NUMBER: 3990033	
PROCEDURE USED: ProSim 8 VER - Exp Unc WW	REPORT NUMBER: 2022005957
TECHNICIAN: James Robinson	ISSUE DATE: 07 SEP 2022
TEMPERATURE: 21.77 °C	CALIBRATION DATE: 07 SEP 2022
HUMIDITY: 55.2 %rh	REQUESTED RECALIBRATION DATE: 07 SEP 2023
SERVICE ORDER NUMBER: 3256358	AS-FOUND STATUS: N/A
	AS-LEFT STATUS: Pass

Fluke Biomedical certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure unless otherwise noted. It has been calibrated using measurement standards traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), or other national metrology institutes.

The Fluke Biomedical quality system complies with ISO/IEC 17025:2017. Although the item calibrated may meet specification and performance criteria at the time of calibration, due to any number of factors the recommended due date of this item does not imply continuing conformance to specifications during the recommended interval.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

This report may contain results of tests for which no NIST calibration standard exists, such as oxygen saturation (SpO2). These tests are conducted using in-house performance standards and accepted test procedures.

Any Test Uncertainty Ratio (TUR) that is less than four-to-one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four-to-one, the field is left blank. Test status is PASS unless noted otherwise.

REMARKS:

APPROVED BY: *Serome Barrow*
 Serome Barrow
 Biomedical Service Manager