

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 12 December Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon	12/1/2025	0835	-22.6	-27.8	-21.3	MW	
Tue	12/2/2025	0814	-21.7	-27.3	-21.3	MW	
Wed	12/3/2025	0815	-22.0	-28.3	-19.8	ADK	
Thu	12/4/2025	0740	-22.4	-24.3	-21.4	MW	
Fri	12/5/2025	0716	-22.5	-26.4	-21.2	MW	
Mon	12/8/2025	0815	-23.0	-27.9	-21.5	ADK	
Tue	12/9/2025	0748	-22.7	-26.9	-21.5	MW	
Wed	12/10/2025	0839	-23.5	-28.5	-21.4	MW	
Thu	12/11/2025	0800	-22.7	-27.3	-21.6	ADK	
Fri	12/12/2025	0820	-22.5	-27.7	-21.4	ADK	
Mon	12/15/2025	0757	-22.5	-25.8	-21.4	MW	
Tue	12/16/2025	0815	-22.9	-28.0	-17.8	ADK	
Wed	12/17/2025	0805	-16.7	-28.3	-13.5	MW	
Thu	12/18/2025	1520	-28.1	-29.9	-17.3	MW	
Fri	12/19/2025	1515	-28.1	-29.9	-17.3	W	
Mon	12/22/2025	0801	-23.3	-29.7	-22.7	MW	
Tue	12/23/2025	0809	-23.3	-29.7	-22.7	MW	
Wed	12/24/2025	0802	-23.4	-29.5	-22.6		
Thu	12/25/2025	N/A	N/A	N/A	N/A	MW	Closed for the holiday
Fri	12/26/2025					W	Closed for the holiday
Mon	12/29/2025					W	Closed for the holiday
Tue	12/30/2025					W	Closed for the holiday
Wed	12/31/2025					W	Closed for the holiday

Key: * Heavy usage