

Service Report



WORK ORDER ID: CWKD4083079

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy, Ste. #1360 Fairway, KS 66205

AGILITI ASSET ID : 1887914 | SERIAL #: N/A | VENDOR SITE ID: | CUSTOMER ASSET ID: VS24847167

ASSET TYPE: REFRIGERATORS, LABORATORY | MANUFACTURER: WHIRLPOOL CORP | MODEL: ED25DQXAN00

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887914

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Kirstin Williams	PM Not Covered	0.5	M-F 8-5	01/23/24

SERVICE PROVIDED DETAILS

Performed SM per OEM specifications. All test passed.

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
QC8	139	SAFETY ANALYZER	04/14/23
DM1	459	DIGITAL MULTIMETER	04/17/23

CHECKLIST

PROCEDURE: L1100 L1

QUESTION	ANSWER
CES THERM. - FREEZER (if applicable)	-13
CES THERM. - REFRIGERATOR	13
CUST. THERM. - FREEZER (if applicable)	-14
CUST. THERM. - REFRIGERATOR	13
DESIRED TEMP. - FREEZER (if applicable)	-15
DESIRED TEMP. - REFRIGERATOR	13
Electrical Safety Inspection	Pass
GASKET SEAL	Pass
High Temperature Alarm	Yes
Low Temperature Alarm	Yes
Mechanical Safety Inspection	Pass
Operational Check	Pass
STANDS LEVEL	Pass
STANDS SOLIDLY	Pass

Customer Signature:

Date:

Technician Signature:

Date: