

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 10 October Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Wed	10/1/2025	0745	-24.0	-28.6	-23.0	MW	
Thu	10/2/2025	0735	-23.9	-28.5	-23.1	MW	
Fri	10/3/2025	0827	-24.7	-28.4	-22.5	MW	
Mon	10/6/2025	0835	-23.5	-28.7	-23.0	ADK	
Tue	10/7/2025	0815	-24.5	-28.7	-17.4	ADK	
Wed	10/8/2025	0849	-24.0	-28.5	-22.6	MW	
Thu	10/9/2025	0750	-25.6	-28.3	-22.5	MW	
Fri	10/10/2025	0830	-23.0	-28.6	-22.6	MW	
Mon	10/13/2025	0810	-23.7	-29.0	-22.5	ADK	
Tue	10/14/2025	0742	-23.2	-29.0	-22.5	MW	
Wed	10/15/2025	0729	-24.9	-28.8	-22.5	MW	
Thu	10/16/2025	0747	-24.6	-27.8	-21.9	MW	
Fri	10/17/2025	0750	-21.8	-27.9	-17.0	MW	
Mon	10/20/2025	0800	-23.0	-28.0	-20.8	ADK	
Tue	10/21/2025	0757	-24.2	-28.0	-21.9	MW	
Wed	10/22/2025	0751	-22.3	-27.9	-21.8	MW	
Thu	10/23/2025	0725	-22.3	-27.9	-21.8	MW	
Fri	10/24/2025	0738	-22.9	-28.8	-21.8	MW	
Mon	10/27/2025	0749	-23.1	-28.2	-21.8	MW	
Tue	10/28/2025	0815	-24.8	-28.3	-22.0	ADK	
Wed	10/29/2025	0805	-23.2	-28.0	-22.0	MW	
Thu	10/30/2025	0815	-22.9	-28.0	-21.8	ADK	
Fri	10/31/2025	0745	-22.2	-28.0	-21.8	MW	

Key: * Heavy usage