

Thermometer Performance Report

Customer:		University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)					Contact:		Andra Lahner					
Street:		3901 Rainbow					Phone #		913-574-3018					
City/ST/Zip:		Kansas City, KS 66160					+/- 2 C=Pass		P= Pass		F=Fail			
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Performed	Date Due
											Unit Temp	Certified Temp		
S-1	TSX60086A	THERMO	1117686301191108	P	P	P	P	P	P	P	-78	-78.1	1/12/2023	1/2024
S-2	ABT-HC-19TS	ABS	ABS-9576207-1907	P	P	P	P	P	P	P	4.5	4.7	1/12/2023	1/2024
S-3	2-F-4	AEGIS	0820872-38728-Y	P	P	P	P	P	P	P	-19.6	-19.6	1/12/2023	1/2024
S-4	ISOTEMP	FISHER	A434F1FF2A93	P	P	P	P	P	P	P	NA	2.6	1/12/2023	1/2024
S-5	GEMD270A	ICE-O-MATIC	Y9 171480 18P								CLEAN CONDENSER			
DELP G040														
1124092101210920	TSX0086D	THERMO	1124092101210920	P	P	P	P	P	P	P	-79	-79.4	1/12/2023	1/2024
	VT65ML	ACCUCOLD	202112000719	P	P	P	P	P	P	P	-24.4	-24.5	1/12/2023	1/2024
	3751FS	THERMO	0146956301121018	P	P	P	P	P	P	P	NA	-15	1/12/2023	1/2024

Technician:

Test equipment used
FLUKE- 52II Thermometer SN: 45470242WS EX: 07/12/2023, #74293

Notes:

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088

Thermometer Performance Report

Customer: University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)		Contact: Andra Lahner												
Street: 3901 Rainbow		Phone #: 913-574-3018												
City/ST/Zip: Kansas City, KS 66160		+/- 2 C=Pass P= Pass F=Fail												
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Peformed	Date Due
											Unit Temp	Certified Temp		

Thermometer Performance Report

Customer: University of Kansas Health Syst-CTSU (Rainbow)(Sudler 2012)		Contact: Andra Lahner												
Street: 3901 Rainbow		Phone #: 913-574-3018												
City/ST/Zip: Kansas City, KS 66160		+/- 2 C=Pass P= Pass F=Fail												
Asset#	Model#	MFG	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Calibration Pass/Fail +/- 2C	Temperature (Celsius)		Date Peformed	Date Due
											Unit Temp	Certified Temp		