

Scale Performance Report

Customer:		KU-CTSU Rainbow				Contact:		Shelby Willams							
Street:		3901 Rainbow Blvd (2nd Floor)				Phone #		913-574-3018							
City/ST/Zip:		Kansas City, KS 66160				P= Pass		F=Fail							
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact							Date Performed
									Target	Actual	Target	Actual	Target	Actual	
Health-O-Meter	500KL	5000163137	P	P	P	P	P	P	50.0	49.8	250.0	250.6	500.0	500.3	01/31/2022
Detecto	MV1	E17619-0049	P	P	P	P	P	P	50.0	49.8	250.0	250.6	500.0	501.4	01/31/2022
Detecto	500KL	5000186170	P	P	P	P	P	P	50.0	49.8	250.0	249.4	500.0	499.2	01/31/2022
DELP G040															
Health-O-Meter	500KL	5000178353	P	P	P	P	P	P	50.0	50.0	250.0	250.2	500.0	500.0	01/31/2022
Health-O-Meter	500KL	5000163141	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	500.0	01/31/2022

Technicians Initials	JC		Technician:	Jim Cowart		Test equipment used	50lb Weights: Kansas Metrology Laboratory Report Number: 15825 Expires: 02/2023	
	Notes:	Next Due for Calibration 01/2023						

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088

