

Service Report



WORK ORDER ID: CWKD4088988

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | **DEPARTMENT:** 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | **ACCOUNT NUMBER:** 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy. Ste. #1360 Fairway, KS 66205

AGILITY ASSET ID : 1485862 | **SERIAL #:** ALZ19A019 | **VENDOR SITE ID :** | **CUSTOMER ASSET ID:**

ASSET TYPE: CENTRIFUGES, TABLETOP, HIGH-SPEED, REFRIGERATED | **MANUFACTURER:** BECKMAN COULTER INC | **MODEL:** ALLEGRA X-30R

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1485862

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

LABOR

| SERVICE PROVIDER | CATEGORY | HOURS | PERIOD DEFINITION | SERVICE DATE |
|-------------------|----------------|-------|-------------------|--------------|
| Marcus Boatwright | PM Not Covered | 0.3 | M-F 8-5 | 01/23/24 |

SERVICE PROVIDED DETAILS

Completed SM passed inspection testing

TEST EQUIPMENT

| PREFIX | UNIT | DESCRIPTION | CALIBRATION DATE |
|--------|------|-------------|------------------|
| TA1 | 257 | TACHOMETER | 04/14/23 |

CHECKLIST

PROCEDURE: CENTRIFUGE

QUESTION

- BRUSHES
- COVER LATCH
- Electrical Safety Inspection
- Mechanical Safety Inspection
- MOTOR CLEANED
- MOTOR CURRENT
- ROTOR & MOTOR
- RPM CHECK
- SAFETY INTERLOCK
- TACHOMETER SERIAL NUMBER
- TEMPERATURE VERIFICATION
- TIMER

ANSWER

- Pass
- Pass
- Pass
- Pass
- NA
- Pass
- Pass
- Pass
- Pass
- 17263217
- NA
- 2800 RPM

Customer Signature:

Date:

Technician Signature:

Date: