

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: CTSU Rainbow – Sudler 2012E

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda (913)547-3007

Thermometer Information: Patient Thermometer

Manufacturer: Covidien

Model: Filac EZ

Serial No: A1952323X

Certified Temperature (°F): 99.14 °F

Observed Temperature (°F): 98.80 °F

Observed Accuracy (°F): 0.34

Passed/Failed: Passed

Certification Date: 02/14/2025

Certification Due: 02/2026

Certifying Equipment

Manufacturer: Fluke

Model: 54 II B

Serial No: 98060021

Due Date: 06/25/2025

Method: Certified per manufacturer's specifications and procedures traceable to NIST

**Service Technician: Andrew Rawitch
(Print)**

Date: 02/14/2025

**Service Technician: *Andrew Rawitch*
(Sign)**

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In

Manufacturer: Coviden

Model: 10011658

Serial No: A1952353X

Certified Temperature (°C): 40.4

Observed Temperature (°C): 41.1

Observed Accuracy (°C): 0.7

Passed/Failed: Pass

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke

Model: 52II

Serial No: 15100368

Due Date: 2/2/2025

Service Technician: Derek Hulett
(Print)

Date: 1/7/2025

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In

Manufacturer: Coviden

Model: 10011658

Serial No: A1952371X

Certified Temperature (°C): 40.5

Observed Temperature (°C): 40.8

Observed Accuracy (°C): 0.3

Passed/Failed: Pass

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke

Model: 52II


Serial No: 15100368

Due Date: 2/2/2025

Service Technician: Derek Hulett
(Print)

Date: 1/7/2025

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In

Manufacturer: WelchAllyn

Model: 901053

Serial No: (21)19147463

Certified Temperature (°C): 40.2

Observed Temperature (°C): 39.4

Observed Accuracy (°C): 0.8

Passed/Failed: Pass

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke

Model: 52II

Serial No: 15100368

Due Date: 2/2/2025

Service Technician: Derek Hulett
(Print)

Date: 1/7/2025

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In

Manufacturer: WelchAllyn

Model: 901053

Serial No: (21)19147495

Certified Temperature (°C): 40.1

Observed Temperature (°C): 40.0

Observed Accuracy (°C): 0.1

Passed/Failed: Pass

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke

Model: 52II


Serial No: 15100368

Due Date: 2/2/2025

Service Technician: Derek Hulett
(Print)

Date: 1/7/2025

Service Technician:
(Sign)



COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: CTSU Rainbow – Sudler 2012E

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda (913)547-3007

Thermometer Information: Patient Thermometer

Manufacturer: Welch Allyn

Model: Sure Temp Plus

Serial No: 04293852

Certified Temperature (°F): 99.14 °F

Observed Temperature (°F): 100.40 °F

Observed Accuracy (°F): 1.26

Passed/Failed: Passed

Certification Date: 02/14/2025

Certification Due: 02/2026

Certifying Equipment

Manufacturer: Fluke

Model: 54 II B

Serial No: 98060021

Due Date: 06/25/2025

Method: Certified per manufacturer's specifications and procedures traceable to NIST

**Service Technician: Andrew Rawitch
(Print)**

Date: 02/14/2025

**Service Technician: *Andrew Rawitch*
(Sign)**

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: CTSU Rainbow – Sudler 2012E

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda (913)547-3007

Thermometer Information: Patient Thermometer

Manufacturer: Welch Allyn

Model: Sure Temp Plus

Serial No: 19147359

Certified Temperature (°F): 99.14 °F

Observed Temperature (°F): 99.00 °F

Observed Accuracy (°F): 0.14

Passed/Failed: Passed

Certification Date: 02/14/2025

Certification Due: 02/2026

Certifying Equipment

Manufacturer: Fluke

Model: 54 II B

Serial No: 98060021

Due Date: 06/25/2025

Method: Certified per manufacturer's specifications and procedures traceable to NIST

**Service Technician: Andrew Rawitch
(Print)**

Date: 02/14/2025

**Service Technician: *Andrew Rawitch*
(Sign)**