

Lab Fridge Monthly
for
CRC 3537 951703280749W

Jan 01, 2024 - Feb 01, 2024

Author: Kristy Anderson
kanderson6@kumc.edu

Review & approval records location:

Summary Report "Lab Fridge Monthly" for CRC 3537 951703280749W

Jan 01, 2024 - Feb 01, 2024

MACHINE DETAILS


Serial Number: **60777160A8E3**
Location: **Unknown**
Description: **60777160A8E3**
Installation date:

EQUIPMENT DETAILS

Category: **Unknown**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **3.73 °C** | MIN: **1.23 °C** | MAX: **11.31 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2023-12-31 6:00 AM	4.08°C
2024-01-01 6:00 AM	2.8°C
2024-01-02 6:00 AM	3.17°C
2024-01-03 6:00 AM	2.16°C
2024-01-04 6:00 AM	5.39°C
2024-01-05 6:00 AM	2.27°C
2024-01-06 6:00 AM	5.35°C
2024-01-07 6:00 AM	3.99°C
2024-01-08 6:00 AM	3.71°C
2024-01-09 6:00 AM	1.48°C
2024-01-10 6:00 AM	3.15°C
2024-01-11 6:00 AM	4.58°C
2024-01-12 6:00 AM	5.34°C
2024-01-13 6:00 AM	2.77°C
2024-01-14 6:00 AM	2.27°C
2024-01-15 6:00 AM	3.73°C
2024-01-16 6:00 AM	4.45°C
2024-01-17 6:00 AM	3.94°C
2024-01-18 6:00 AM	4.95°C
2024-01-19 6:00 AM	5.74°C
2024-01-20 6:00 AM	2.49°C
2024-01-21 6:00 AM	3.03°C
2024-01-22 6:00 AM	3.67°C
2024-01-23 6:00 AM	3.36°C
2024-01-24 6:00 AM	4.87°C
2024-01-25 6:00 AM	3.4°C
2024-01-26 6:00 AM	3.27°C
2024-01-27 6:00 AM	2.23°C
2024-01-28 6:00 AM	3.09°C
2024-01-29 6:00 AM	4.16°C
2024-01-30 6:00 AM	3.15°C

CENTRAL TIME (US & CANADA)

TEMPERATURE (°C)





2024-01-31 6:00 AM

2.98°C

THRESHOLD ALERTS

No Alert Events for selected date range.

CONNECTION ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	ACTIVITY
2024-01-25 6:34 PM	CRC 3537 951703280749W	Connection	Connection Alert	Connected	Alert Cleared
	2024-01-25 6:34 PM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W has restored wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-01-25 6:34 PM	Message:CRC 3537 951703280749W has restored wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-01-25 6:32 PM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W has lost wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-01-25 6:32 PM	Message:CRC 3537 951703280749W has lost wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Lab Fridge Monthly
for
CRC 3537 951703280749W

Feb 01, 2024 - Mar 01, 2024

Author: Kristy Anderson
kanderson6@kumc.edu

Review & approval records location:

Summary Report "Lab Fridge Monthly" for CRC 3537 951703280749W

Feb 01, 2024 - Mar 01, 2024

MACHINE DETAILS


Serial Number: **60777160A8E3**
Location: **Unknown**
Description: **60777160A8E3**
Installation date:

EQUIPMENT DETAILS

Category: **Unknown**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **3.73 °C** | MIN: **1.23 °C** | MAX: **10.82 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-01-31 6:00 AM	3.63°C
2024-02-01 6:00 AM	2.67°C
2024-02-02 6:00 AM	3.48°C
2024-02-03 6:00 AM	2.31°C
2024-02-04 6:00 AM	2.12°C
2024-02-05 6:00 AM	1.57°C
2024-02-06 6:00 AM	1.91°C
2024-02-07 6:00 AM	2.86°C
2024-02-08 6:00 AM	3.44°C
2024-02-09 6:00 AM	3.57°C
2024-02-10 6:00 AM	4.27°C
2024-02-11 6:00 AM	3.4°C
2024-02-12 6:00 AM	3.73°C
2024-02-13 6:00 AM	3.28°C
2024-02-14 6:00 AM	1.9°C
2024-02-15 6:00 AM	5.55°C
2024-02-16 6:00 AM	4.27°C
2024-02-17 6:00 AM	2.8°C
2024-02-18 6:00 AM	5.22°C
2024-02-19 6:00 AM	5.29°C
2024-02-20 6:00 AM	3.69°C
2024-02-21 6:00 AM	2.51°C
2024-02-22 6:00 AM	4.74°C
2024-02-23 6:00 AM	1.79°C
2024-02-24 6:00 AM	1.61°C
2024-02-25 6:00 AM	2.04°C
2024-02-26 6:00 AM	5.52°C
2024-02-27 6:00 AM	5.34°C
2024-02-28 6:00 AM	1.9°C
2024-02-29 6:00 AM	3.6°C

ALERTS

No Alert Events for selected date range.

Lab Fridge Monthly
for
CRC 3537 951703280749W

Mar 01, 2024 - Apr 01, 2024

Author: Kristy Anderson
kanderson6@kumc.edu

Review & approval records location:

Summary Report "Lab Fridge Monthly" for CRC 3537 951703280749W

Mar 01, 2024 - Apr 01, 2024

MACHINE DETAILS


Serial Number: **60777160A8E3**
Location: **Unknown**
Description: **60777160A8E3**
Installation date:

EQUIPMENT DETAILS

Category: **Unknown**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **3.72 °C** | MIN: **1.26 °C** | MAX: **9.48 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-02-29 5:00 AM	4.91°C
2024-03-01 5:00 AM	5.59°C
2024-03-02 5:00 AM	5.41°C
2024-03-03 5:00 AM	1.76°C
2024-03-04 5:00 AM	3.89°C
2024-03-05 5:00 AM	5.58°C
2024-03-06 5:00 AM	4.02°C
2024-03-07 5:00 AM	2.91°C
2024-03-08 5:00 AM	3.69°C
2024-03-09 5:00 AM	2.6°C
2024-03-10 5:00 AM	2.35°C
2024-03-11 5:00 AM	4.73°C
2024-03-12 5:00 AM	2.52°C
2024-03-13 5:00 AM	3.69°C
2024-03-14 5:00 AM	2.93°C
2024-03-15 5:00 AM	4.64°C
2024-03-16 5:00 AM	5.02°C
2024-03-17 5:00 AM	4.97°C
2024-03-18 5:00 AM	3.73°C
2024-03-19 5:00 AM	5.02°C
2024-03-20 5:00 AM	2.58°C
2024-03-21 5:00 AM	4.28°C
2024-03-22 5:00 AM	4.39°C
2024-03-23 5:00 AM	4.12°C
2024-03-24 5:00 AM	5.2°C
2024-03-25 5:00 AM	4.89°C
2024-03-26 5:00 AM	3.55°C
2024-03-27 5:00 AM	3.02°C
2024-03-28 5:00 AM	5.22°C
2024-03-29 5:00 AM	4.09°C
2024-03-30 5:00 AM	2.34°C

CENTRAL TIME (US & CANADA)

TEMPERATURE (°C)





2024-03-31 5:00 AM

2.93°C

THRESHOLD ALERTS

No Alert Events for selected date range.

CONNECTION ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	ACTIVITY
2024-03-06 2:24 PM	CRC 3537 951703280749W	Connection	Connection Alert	Connected	Alert Cleared
	2024-03-06 2:24 PM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W has restored wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-03-06 2:24 PM	Message:CRC 3537 951703280749W has restored wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-03-06 2:19 PM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W has lost wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			
	2024-03-06 2:19 PM	Message:CRC 3537 951703280749W has lost wireless connection to the Elemental Gateway. Alert Rule:Connection Alert			

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Lab Fridge Monthly
for
CRC 3537 951703280749W

Apr 01, 2024 - May 01, 2024

Author: Kristy Anderson
kanderson6@kumc.edu

Review & approval records location:

Summary Report "Lab Fridge Monthly" for CRC 3537 951703280749W

Apr 01, 2024 - May 01, 2024


MACHINE DETAILS

Serial Number: **60777160A8E3**
Location: **Unknown**
Description: **60777160A8E3**
Installation date:

EQUIPMENT DETAILS

Category: **Unknown**
Manufacturer:
Model:
Service Entered Date:





TEMPERATURE AVG: **3.71 °C** | MIN: **1.24 °C** | MAX: **15.9 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-03-31 5:00 AM	2.88°C
2024-04-01 5:00 AM	2.22°C
2024-04-02 5:00 AM	4.37°C
2024-04-03 5:00 AM	5.02°C
2024-04-04 5:00 AM	4.74°C
2024-04-05 5:00 AM	1.55°C
2024-04-06 5:00 AM	3.46°C
2024-04-07 5:00 AM	1.75°C
2024-04-08 5:00 AM	5.08°C
2024-04-09 5:00 AM	3.98°C
2024-04-10 5:00 AM	2.92°C
2024-04-11 5:00 AM	3.77°C
2024-04-12 5:00 AM	2.7°C
2024-04-13 5:00 AM	3.77°C
2024-04-14 5:00 AM	4.66°C
2024-04-15 5:00 AM	5.2°C
2024-04-16 5:00 AM	2.74°C
2024-04-17 5:00 AM	3.44°C
2024-04-18 5:00 AM	5.28°C
2024-04-19 5:00 AM	2.77°C
2024-04-20 5:00 AM	2.66°C
2024-04-21 5:00 AM	4.7°C
2024-04-22 5:00 AM	4.08°C
2024-04-23 5:00 AM	4.04°C
2024-04-24 5:00 AM	2.77°C
2024-04-25 5:00 AM	1.56°C
2024-04-26 5:00 AM	3.01°C
2024-04-27 5:00 AM	4.91°C
2024-04-28 5:00 AM	2.22°C
2024-04-29 5:00 AM	3.62°C
2024-04-30 5:00 AM	3.53°C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-04-11 8:50 AM	CRC 3537 951703280749W	Temperature	Refrigerator Alert	5.95 °C	10.0/1.0 °C	Alert Cleared
	2024-04-11 8:50 AM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W is reporting Temperature of 5.95 °C, within the limit of 1.0 °C to 10.0 °C. Alert Rule:Refrigerator Alert				
	2024-04-11 8:50 AM	Message:Temperature Clear: 5.95 °C, within the limit of 1.0 °C to 10.0 °C Alert Rule:Refrigerator Alert				
	2024-04-11 8:00 AM	Notification:Email and SMS sent to Miranda Wright, Kristy Anderson. Message:CRC 3537 951703280749W is reporting Temperature of 13.59 °C, above the limit of 10.0 °C. Alert Rule:Refrigerator Alert				
	2024-04-11 8:00 AM	Message:Temperature Alert: 13.59 °C, above the limit of 10.0 °C Alert Rule:Refrigerator Alert				

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.