

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Thermo Fisher Scientific

Model: IVD

Serial No: 42665624

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: x	Fail: _____
Lid Latch Safety Interlock	Pass: x	Fail: _____
Brake	Pass: x	Fail: _____
Speed Check Low	Set: 300	Measured: 300.9
Speed Check Med	Set: 2500	Measured: 2499
Speed Check High	Set: 5000	Measured: 4999

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer:

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Beckman-Coulter

Model: AllegraX-30R

Serial No: A1219F037

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: x	Fail: _____
Lid Latch Safety Interlock	Pass: x	Fail: _____
Brake	Pass: x	Fail: _____
Speed Check Low	Set: 100	Measured: 100.3
Speed Check Med	Set: 2300	Measured: 2299
Speed Check High	Set: 4700	Measured: 4700
Temperature	Set: 3	Measured: 3.5

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke 52II SN 15100368 Due 2/2/25

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Beckman-Coulter

Model: AllegraX-30R

Serial No: A1219P0508

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: x	Fail: _____
Lid Latch Safety Interlock	Pass: x	Fail: _____
Brake	Pass: x	Fail: _____
Speed Check Low	Set: 100	Measured: 99.5
Speed Check Med	Set: 2300	Measured: 2299
Speed Check High	Set: 4700	Measured: 4699
Temperature	Set: 22	Measured: 20.9

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke 52II SN 15100368 Due 2/2/25

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Beckman-Coulter

Model: AllegraX-30R

Serial No: ALZ19E044

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: x	Fail: _____
Lid Latch Safety Interlock	Pass: x	Fail: _____
Brake	Pass: x	Fail: _____
Speed Check Low	Set: 100	Measured: 99.6
Speed Check Med	Set: 2300	Measured: 2300
Speed Check High	Set: 4700	Measured: 4699
Temperature	Set: 3	Measured: 1.5

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke 52II SN 15100368 Due 2/2/25

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Beckman-Coulter

Model: Z2

Serial No: 2291H

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: <u> x </u>	Fail: <u> </u>
Lid Latch Safety Interlock	Pass: <u> x </u>	Fail: <u> </u>
Brake	Pass: <u> x </u>	Fail: <u> </u>
Speed Check Low	Set: <u> 100 </u>	Measured: <u> 99.6 </u>
Speed Check Med	Set: <u> 2200 </u>	Measured: <u> 2202 </u>
Speed Check High	Set: <u> 4400 </u>	Measured: <u> 4388 </u>
Temperature	Set: <u> 18 </u>	Measured: <u> 18.7 </u>

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke 52II SN 15100368 Due 2/2/25

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Equipment Information:

Manufacturer: Beckman-Coulter

Model: AllegraX-30R

Serial No: ALZ20L068

Physical Condition:

Chassis/ Housing:	Pass	X	Fail
Controls & Switches:	Pass	X	Fail
Power Cord & Strain Relief:	Pass	X	Fail
Hardware Integrity:	Pass	X	Fail
Indicators & Display Operation:	Pass	X	Fail
Mfg. Labeling Intact:	Pass	X	Fail
Condenser Filter Replaced:	Pass	X	Fail

Operation:

Timer Accuracy	Pass: x	Fail: _____
Lid Latch Safety Interlock	Pass: x	Fail: _____
Brake	Pass: x	Fail: _____
Speed Check Low	Set: 100	Measured: 99.9
Speed Check Med	Set: 2300	Measured: 2300
Speed Check High	Set: 4700	Measured: 4700
Temperature	Set: 3	Measured: 2.6

Certification Date: 1/7/2025

Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke 52II SN 15100368 Due 2/2/25

Manufacturer: Extech Photo Tachometer SN H.437785 Due 3/19/25

Manufacturer:

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Derek Hulett

Date: 1/7/25

Service Technician: 