

Service Report



WORK ORDER ID: CWKD4070914

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160 | **SHIPPING ADDRESS:** 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITI ASSET ID : 1887878 | **SERIAL #:** 11922000794 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: ELECTROCARDIOGRAPHS, MULTICHANNEL, INTERPRETIVE | **MANUFACTURER:** WELCH ALLYN INC A SUBSIDIARY OF HILLROM HOLDINGS INC | **MODEL:** ELI 280

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887878

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-19-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Marcus Boatwright	PM Not Covered	0.22	M-F 8-5	01/19/24
Stephanie Harper	PM Not Covered	0.22	M-F 8-5	01/19/24

SERVICE PROVIDED DETAILS

Completed SM passed all inspection testing 102/42 ES

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
PR3	473	OXIMETER SIMULATOR	04/14/23

CHECKLIST

PROCEDURE: M1230 L1

QUESTION	ANSWER
CABLES	Pass
CALIBRATION	Pass
CLEAN	Yes
CLEAN	Yes
CLEAN/LUBE REC.	Yes
Clinical- Audio	Pass
Clinical- Visual	Pass
CONNECTORS	Pass
Electrical Safety Inspection	Pass
EXT CONDITION	Pass
MECH	Pass
Mechanical Safety Inspection	Pass
Operational Check	Pass
Patient- Audio	Pass
Patient- Visual	Pass
PLUGS	Pass
Remote	Pass
Safety- Audio	Pass
Safety- Visual	Pass
SPEED	Pass
STYLUS/PRINTHEA	Yes
WAVEFORM	Pass

Customer Signature:

Date:

Technician Signature:

Date: