

CTSU Lab Freezer Min/Max Temperature Log

Maintain temperature between -65°C to -90 °C

Month **February**

Year **2024**

Freezer ID/Location: Lab room freezer 412269-86 (-80 freezer B)

Date (Weekdays only)	Time	Min °C	Max °C	Initials	Comments
Thu 2/1/2024	0900	-75.06	-70.06	DN	
Fri 2/2/2024	0900	-76.19	5.78	DN	*
Mon 2/5/2024	0800	-76.27	-73.45	DN	
Tue 2/6/2024	0800	-76.25	-64.67	LM	
Wed 2/7/2024	0800	-76.63	-44.48	DN	*
Thu 2/8/2024	0800	-76.06	-50.91	DN	*
Fri 2/9/2024	0800	-76.06	-61.81	DN	*
Mon 2/12/2024	0900	-75.61	-73.20	DN	
Tue 2/13/2024	0900	-75.98	-62.53	DN	*
Wed 2/14/2024	0800	-76.03	-52.53	LM	*
Thu 2/15/2024	0900	-76.03	-56.44	DN	*
Fri 2/16/2024	0800	-75.94	-52.20	LM	*
Mon 2/19/2024	0000	-76.14	-70.21	DN	
Tue 2/20/2024	0915	-76.14	-52.23	DN	*
Wed 2/21/2024	0900	-75.94	-47.11	DN	*
Thu 2/22/2024	1000	-75.92	-63.66	DN	
Fri 2/23/2024	1000	-75.97	-61.80	DN	
Mon 2/26/2024	0800	-76.41	-73.77	DN	
Tue 2/27/2024	0830	-76.41	-62.76	DN	%, *
Wed 2/28/2024	1000	-76.14	-62.76	DN	
Thu 2/29/2024	1000	-75.78	-65.47	DN	

Key: * Heavy usage
 % Normal peak of cooling curve
 # Normal trough of cooling curve