

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 35°C to 37°C

Incubator - Sudler

Boekel 192423685

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month **5** May Year **2025** Threshold: **30 minutes**

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Thu	5/1/2025	0715	37.4	36.8	37.9	MW	
Fri	5/2/2025	0735	37.7	34.9	38.4	MW	
Mon	5/5/2025	0715	37.4	36.8	38.8	MW	
Tue	5/6/2025	0705	37.4	36.8	38.8	MW	
Wed	5/7/2025	0826	38.4	37.2	39.1	MW	
Thu	5/8/2025	0737	38.0	36.2	38.5	MW	
Fri	5/9/2025	0715	37.1	36.2	38.5	MW	
Mon	5/12/2025	0810	35.9	35.9	38.4	MW	
Tue	5/13/2025	0828	36.7	35.9	38.6	MW	
Wed	5/14/2025	0715	36.7	38.5	38.4	MW	
Thu	5/15/2025	0720	38.3	36.7	40.1	MW	
Fri	5/16/2025	0727	37.2	36.7	38.3	MW	
Mon	5/19/2025	0710	37.2	36.7	38.3	MW	
Tue	5/20/2025	0713	37.3	36.8	38.9	MW	
Wed	5/21/2025	0725	38.0	33.8	38.4	MW	
Thu	5/22/2025	0800	37.8	36.8	38.2	ADK	
Fri	5/23/2025	0800	37.8	36.7	38.4	ADK	
Mon	5/26/2025						Holiday - DH
Tue	5/27/2025	0800	37.8	35.9	38.2	ADK	
Wed	5/28/2025	1130	36.9	36	38.0	DH	
Thu	5/29/2025	0800	37.0	36.4	38.0	DH	
Fri	5/30/2025	0800	37.7	36.4	38.0	DH	
Mon	6/2/2025						

Key: * Heavy usage