

Service Report



WORK ORDER ID: CWKD4081779

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: KUMC RESEARCH INSTITUTE | DEPARTMENT: 190215-KUMC RESEARCH INSTITUTE-ROUNDS RESEARCH INSTITUTE | ACCOUNT NUMBER: 190215

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Kansas City, KS 66103

SHIPPING ADDRESS: 4330 Shawnee Mission Pkwy. Ste. #1360 Fairway, KS 66205

AGILITI ASSET ID : 1887933 | SERIAL #: NONE | VENDOR SITE ID: | CUSTOMER ASSET ID:

ASSET TYPE: REFRIGERATORS, LABORATORY | MANUFACTURER: FRIGIDAIRE | MODEL: 61441

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1887933

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-23-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Harper	PM Not Covered	0.3	M-F 8-5	01/23/24

SERVICE PROVIDED DETAILS

Completed SM passed inspection testing

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
DM1	286	DIGITAL MULTIMETER	04/14/23

CHECKLIST

PROCEDURE: L1100 L1

QUESTION

- CES THERM. - FREEZER (if applicable)
- CES THERM. - REFRIGERATOR
- CUST. THERM. - FREEZER (if applicable)
- CUST. THERM. - REFRIGERATOR
- DESIRED TEMP. - FREEZER (if applicable)
- DESIRED TEMP. - REFRIGERATOR
- Electrical Safety Inspection
- GASKET SEAL
- High Temperature Alarm
- Low Temperature Alarm
- Mechanical Safety Inspection
- Operational Check
- STANDS LEVEL
- STANDS SOLIDLY

ANSWER

- N/A
- N/A
- 19C
- 20C
- N/A
- N/A
- Pass
- Pass
- Yes
- Yes
- Pass
- Pass
- Pass
- Pass

Customer Signature:

Date:

Technician Signature:

Date: