

Scale Performance Report

Customer: UKMC-Swope CTSU		Contact: Shelby Williams														
Street: 4000 Dr. Martin Luther King Jr. Blvd		Phone #: 816-599-5310														
City/ST/Zip: Kansas City, MO 64130		P= Pass F=Fail														
Unit	Model#	MFG	SIN	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strain Relief	Hardware Integrity	Indicators & Display Operation	MFG. Labeling Intact	Target	Actual	Target	Actual	Target	Actual	Date Performed
Floor Scale	PS5700	BEFOUR	1506PB57813	P	P	P	P	P	Y	50	50.2	250	250	500	501.2	11/14/2022
Floor Scale	PS5700	BEFOUR	1601PB57847	P	P	P	P	P	Y	50	49.6	250	249.2	500	499.8	11/14/2022
Floor Scale	PS5700	BEFOUR	0705PS57384	P	P	P	P	P	Y	50	49.8	250	249.3	500	499.8	11/14/2022
Floor Scale	500KL	Pelstar	203362	P	P	P	P	P	Y	50	50.2	250	250.4	500	500.4	3/3/2023

Technicians
Initials

JC

Technician: Jim Cowart

Test equipment used
50lb Weights: Kansas Metrology Laboratory
Report Number: K16669 Expires: 04/1/2023

Notes: Next Due for Calibration 11/2023