

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Aegis Freezer - Sudler

0820872-38728

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 10 October Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Wed	10/1/2025	0722	-21.8	-27.0	-17.7	MW	
Thu	10/2/2025	0800	-25.1	-27.0	-17.7	MW	
Fri	10/3/2025	0755	-18.1	-27.4	-17.5	MW	
Mon	10/6/2025	0800	-23.4	-27.8	-16.1	ADK	
Tue	10/7/2025	0800	-17.7	-28.3	-17.4	ADK	
Wed	10/8/2025	0715	-20.8	-28.4	-17.4	MW	
Thu	10/9/2025	0800	-22.6	-28.4	-17.4	MW	
Fri	10/10/2025	0720	-18.3	-28.1	-16.8	MW	
Mon	10/13/2025	0756	-26.0	-28.1	-16.8	MW	
Tue	10/14/2025	0715	-25.7	-28.0	-16.1	MW	
Wed	10/15/2025	0755	-26.0	-28.0	-16.4	MW	
Thu	10/16/2025	0715	-26.0	-28.0	-16.4	MW	
Fri	10/17/2025	0715	-26.0	-27.9	-17.0	MW	
Mon	10/20/2025	0718	-20.0	-27.9	-17.0	MW	
Tue	10/21/2025	0714	-20.0	-27.9	-17.0	MW	
Wed	10/22/2025	0735	-17.7	-27.8	-17.2	MW	
Thu	10/23/2025	0745	-18.6	-27.8	-17.2	MW	
Fri	10/24/2025	0745	-20.3	-27.9	-17.2	MW	
Mon	10/27/2025	0714	-22.1	-27.8	-16.6	MW	
Tue	10/28/2025	0734	-20.2	-27.4	-16.9	MW	
Wed	10/29/2025	0715	-19.6	-27.3	-17.1	MW	
Thu	10/30/2025	1000	-19.6	-27.3	-17.0	MW	
Fri	10/31/2025	0730	-17.4	-27.5	-17.0	MW	

Key: * Heavy usage