

Service Report



WORK ORDER ID: CWKD6982665

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIVERSITY OF KS MED CTR-CTSU | **DEPARTMENT:** 146148-UNIVERSITY OF KS MED CTR-CTSU-BIOMED | **ACCOUNT NUMBER:** 146148

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Mail Stop 6011 Kansas City, KS 66103

SHIPPING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

AGILITY ASSET ID : 1887905 | **SERIAL #:** 20468939 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: THERMOMETERS, ELECTRONIC | **MANUFACTURER:** WELCH ALLYN INC | **MODEL:** SURETEMP PLUS 690

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-22-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.3	M-F 8-5	01/22/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

CHECKLIST

PROCEDURE: WELCH ALLYN SURETEMP PLUS 690 - ANNUAL

QUESTION

Perform the Power-On Self Test.

ANSWER

Pass

Perform the Model 9600 Plus Calibration Testing procedure. (thermometer's temperature reading and the Tester's Set Point Temperature reading are

within $\pm 0.2^\circ$ F for each Set Point temperature)
Calibration Testing. Cal-Key Verification - The temperature should read 97.3 +/- 0.2°F.

Pass
97.3

Customer Signature:

Date:

Stephanie Dye

1/22/25

Technician Signature:

Date: