

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

DW-304678
 7075759 (-20) BEBYH
 Refrigerator or Freezer ID/Location: SE006
 CTSU FAIRWAY 00JQ9
 -DW 26-11-2025 40008

Month 6 June Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon	6/2/2025	1000	-29	-29	-29	DW	
Tue	6/3/2025	0900	-27	-29	-27	DW	
Wed	6/4/2025	0900	-28	-29	-29	DW	
Thu	6/5/2025	1000	-28	-29	-26	DW	
Fri	6/6/2025	1000	-28	-29	-26	DW	
Mon	6/9/2025	0900	-28	-29	-22	DW	
Tue	6/10/2025	0800	-26	-28	-20	DW	
Wed	6/11/2025	0800	-26	-28	-25	DW	
Thu	6/12/2025	0800	-26	-29	-25	DW	
Fri	6/13/2025	0900	-27.16	-28.70	-24.78	KLA	
Mon	6/16/2025	0900	-27.16	-28.48	-24.66	KLA	
Tue	6/17/2025	1000	-27	-28	-24	DW	
Wed	6/18/2025	1000	-27	-28	-24	DW	
Thu	6/19/2025						Holiday
Fri	6/20/2025	0900	-25.1	-27	-24	DW	
Mon	6/23/2025	1000	-28	-29	-22	DW	
Tue	6/24/2025	1005	-27	-29	-22	KLA	
Wed	6/25/2025	0930	-27	-29	-22	KLA	
Thu	6/26/2025	0900	-28	-29	-22	KLA	
Fri	6/27/2025	0817	-28	-29	-22	KLA	
Mon	6/30/2025	1100	-27	-28	-27	KLA	
Tue	7/1/2025						
Wed	7/2/2025						See July log. DW 11-26-2025

Key: * Heavy usage